**Innovation in low-cost video laryngoscopy:
Intubator V1-Indirect compared with
Storz C-MAC in a simulated difficult airway**

*Minor dissertation submitted in partial fulfilment of the requirements for the
degree of Master of Medicine (MMed) in the
Department of Anaesthesia & Perioperative Medicine*

Dr. Christiaan Tertius de Villiers, MBChB (Pret), DA (SA)

DVLCHR004

Submitted July 2021

Principal supervisor:

Associate Professor Ross Hofmeyr

MBChB (Stell), DipPEC (SA), DA (SA), FCA (SA), MMED (UCT), FAWM

Department of Anaesthesia & Perioperative Medicine

The copyright of this thesis vests in the author. No quotation from it or information derived from it is to be published without full acknowledgement of the source. The thesis is to be used for private study or non-commercial research purposes only.

Published by the University of Cape Town (UCT) in terms of the non-exclusive license granted to UCT by the author.

Declaration page

Plagiarism declaration:

- 1) I am aware that plagiarism is a form of academic dishonesty
- 2) I have read and understood the aspects regarding plagiarism and have avoided all types of plagiarism described in the plagiarism documents.
- 3) I have referenced all quotations and acknowledged the ideas of others.
- 4) I declare that this is my own work and I will not allow others to plagiarise my work.
- 5) I have attached a summary of the “Turnitin” originality report to validate the authenticity of this work.

I, Christiaan de Villiers, hereby declare that the work on which this dissertation/thesis is based is my original work (except where acknowledgements indicate otherwise) and that neither the whole work nor any part of it has been, is being, or is to be submitted for another degree in this or any other university.

I empower the university to reproduce for the purpose of research either the whole or any portion of the contents in any manner whatsoever.

Signature:

Signed by candidate

Date: 5 July 2021

Format

The contents of this minor dissertation are presented in the “*accepted for publication*” format.

This manuscript has been accepted for publication in *Trends in Anaesthesia and Critical Care*.

This work was previously presented at the World Airway Management Meeting (WAMM) in Amsterdam, the Netherlands, for which the abstract of this work was published in *Trends in Anaesthesia and Critical Care*.

Acknowledgements and contributions

Primary author and MMed candidate:

de Villiers CT, MBChB, DA (SA)

Principal supervisor:

Ross Hofmeyr, MBChB, DipPEC, DA (SA), MMed (UCT), FCA (SA), FAWM

Co-Supervisors:

Christella Alphonsus, MBChB, DA (SA), FCA (SA)

Contributors:

Dylan Eave, MBChB, DA (SA)

Dinell Behari, MBChB, DA (SA)

The authors would like to thank their colleagues in the Department of Anaesthesia and Perioperative Medicine of the University of Cape Town for participating in this study. The authors would also like to thank Karl Storz for providing a C-MAC device to conduct this study.

Contents

Declaration page.....	2
Format.....	3
Acknowledgements and contributions.....	4
Abstract.....	7
Chapter 1: Introduction	8
Chapter 2: Manuscript - Accepted for Publication	9
Abstract.....	10
Introduction	11
Methods.....	12
Results.....	13
Discussion	13
Acknowledgements.....	15
Competing interests.....	15
References	16
Chapter 3 – Published Peer-Reviewed Article	21
List of appendices	27
Appendix 1: TACC Acceptance for Publication	28
Appendix 2: TACC Reviewer Comments	29
Appendix 3: Abstract Publication TACC	38
Appendix 4: Human Research Ethics Committee approval	39
Appendix 5: Mannikin study Consent form	40
CONSENT FORM.....	40
Appendix 6: Participant data capture form	41
PARTICIPANT DATA CAPTURE FORM.....	41
Appendix 7: Guide to authors TACC	42

List of tables

Table 1: Intubation time between CMAC and Intubator V1 indirect video laryngoscope

Table 2: Fremantle videolaryngoscopic score for view obtained with CMAC and Intubator

List of figures

Figure 1: V1 Intubator Indirect - Novel low-cost hyperangulated disposable videolaryngoscope

Figure 2: Storz CMAC with Dörge's blade (Karl Storz GmbH, Tuttlingen, Germany)

Figure 3. Equivalence plot showing mean difference in intubation time.

List of appendices

Appendix 1: TACC Acceptance for Publication

Appendix 2: TACC Reviewer Comments

Appendix 3: TACC Abstract Publication

Appendix 4: Human Research Ethics Committee approvals

Appendix 5: Mannikin study consent form

Appendix 6: Participant data capture form

Appendix 7: TACC Guide to authors

Abstract

Background: Video laryngoscopy has directly impacted airway management, with numerous studies demonstrating its utility in clinical management of anatomically difficult airways. However, availability of video laryngoscopes in all clinical areas has been limited by cost. We used smartphone technology, miniature cameras and three-dimensional printing to design and create an innovative low-cost hyperangulated video laryngoscope. This has the potential to make the technique more widely available.

Objectives: The aim of this study was to determine if time to intubation with the novel device was clinically equivalent to an existing gold-standard video laryngoscope (Storz CMAC with Dörge's blade).

Methods: We conducted a randomised, controlled, cross-over equivalence study with 100 skilled practitioners who had previous video laryngoscopy experience. Participants received instruction on the new device, and adequate opportunity to practice. Intubations were then performed in a randomised order on a manikin simulating a difficult airway. Video recordings of each intubation were analysed by two independent investigators to determine time to intubation. A mean difference in intubation time of less than 10 seconds was determined a priori to denote clinical equivalence.

Results: Mean difference in intubation time between the devices was 4.92 seconds, (two one-sided test 95%CI: 2.34 – 7.49 seconds). The innovative low-cost VL was thus clinically equivalent to the industry standard in a simulated difficult airway. Further testing in vivo in a clinical environment is needed.

Conclusion: The results of this study show that a low-cost disposable hyperangulated video laryngoscope is clinically equivalent to the industry standard in a simulated difficult airway. In the context of the current global pandemic, video laryngoscopy has been advised in nearly all airway guidelines. Access to a low-cost VL which does not require reprocessing may be of great value.

Chapter 1: Introduction

Video laryngoscopy has directly impacted airway management, with numerous studies demonstrating its utility in clinical management of anatomically difficult airways. However, availability of video laryngoscopes in all clinical areas has been limited by the relatively high cost. We used smartphone technology, miniature cameras and three-dimensional printing to design and create an innovative low-cost hyperangulated video laryngoscope with the potential to make the technique more widely available.

The aim of this study was to determine if time to intubation with the novel device was clinically equivalent to an existing gold-standard video laryngoscope (Storz CMAC with Dörge blade). We conducted a randomised, controlled, cross-over equivalence study, during April and August 2019 with 100 skilled practitioners who had previous video laryngoscopy experience to determine mean difference in intubation time between the devices.

Chapter 2: Manuscript - Accepted for Publication

The manuscript has been accepted for publication in *Trends in Anaesthesia and Critical Care*.

Submitting author: Christiaan de Villiers

Department of Anaesthesia and Perioperative Medicine
University of Cape Town
D23, Groote Schuur Hospital, Main Road, Cape Town, South Africa

Innovation in low-cost video-laryngoscopy: Intubator V1-Indirect compared with Storz C-MAC in a simulated difficult airway

C. T. de Villiers,¹ C. Alphonsus,² D. Eave,¹ R. Hofmeyr³

1 Senior Registrar,
2 Consultant Anaesthesiologist,
3 Associate Professor

Department of Anaesthesia and Perioperative Medicine,
University of Cape Town,
and Groote Schuur Hospital,
Cape Town, South Africa

Correspondance to: C. T. de Villiers
Email: christiaandevilliers@gmail.com

Keywords: videolaryngoscopy, video laryngoscope, difficult airway, low-cost

Twitter: @rosshofmeyr

This study was undertaken in partial completion of the MMed degree for the first author. Findings of this study have been presented in oral and poster format at the World Airway Management Meeting in Amsterdam, November 2019, and at the South African Society of Anaesthesiologists' annual congress in Pretoria, March 2020. The abstract for WAMM was published in *Trends in Anaesthesia and Critical Care* 2020(30):e127.

Innovation in low-cost video laryngoscopy: Intubator V1-Indirect compared with Storz C-MAC in a simulated difficult airway

Abstract

Background

Video laryngoscopy has directly impacted airway management, with numerous studies demonstrating its utility in clinical management of anatomically difficult airways. However, availability of video laryngoscopes in all clinical areas has been limited by the relatively high cost. We used smartphone technology, miniature cameras and three-dimensional printing to design and create an innovative low-cost hyperangulated video laryngoscope. This has the potential to make the technique more widely available. The aim of this study was to determine if time to intubation with the novel device was clinically equivalent to an existing gold-standard video laryngoscope (Storz CMAC with Dörge blade).

Methods:

We conducted a randomised, controlled, cross-over equivalence study with 100 skilled practitioners who had previous video laryngoscopy experience. Participants received instruction on the new device, and adequate opportunity to practice. Intubations were then performed in a randomised order on a manikin simulating a difficult airway. Video recordings of each intubation were analysed by two independent investigators to determine time to intubation. A mean difference in intubation time of less than 10 seconds was determined *a priori* to denote clinical equivalence.

Results:

Mean difference in intubation time between the devices was 4.92 seconds, (two one-sided test 95%CI: 2.34 – 7.49 seconds). The innovative low-cost VL was thus clinically equivalent to the industry standard in a simulated difficult airway. Further testing *in vivo* in a clinical environment is needed.

Conclusion:

The results of this study show that a low-cost disposable hyperangulated video laryngoscope is clinically equivalent to the industry standard in a simulated difficult airway. In the context of the current global pandemic, video laryngoscopy has been advised in nearly all airway guidelines. Access to a low-cost VL which does not require reprocessing may be of great value.

Introduction

Traditionally, the most commonly used tool for manipulation of the airway and tracheal intubation under direct vision is a Macintosh blade laryngoscope.¹ The ability to safely perform this procedure requires practice and experience, and can prove challenging where there are anatomical or situational factors associated with a difficult airway.²⁻⁵

In human trials, VLs have been shown to improve the view of the larynx and glottis compared to direct laryngoscopy (DL).⁶⁻⁷ VLs have improved intubation first-pass success rates, decreased rates of oesophageal intubation, and increased overall intubation success for both expert and inexperienced providers, within as well as outside the operating theatre context.⁶⁻⁹ Whilst DL intubation has been shown in some studies to be faster than VL, especially in experienced hands, VL may also be an important rescue technique for failed DL intubation. This is reflected in a study which showed the efficacy of the Storz CMAC video laryngoscope as a rescue device in the unanticipated difficult airway, where 86% first-pass success (FPS) was achieved after failed direct laryngoscopy.¹⁰ The Difficult Airway Society (DAS) now recommends that video laryngoscopy should be immediately available whenever intubation is performed.¹¹ A United Kingdom national survey highlighted the fact that VLs are available in 93% of operating theatres, but only 54% of ICUs and 35% of emergency departments.¹¹ However, ICUs and emergency departments are clinical areas where unanticipated difficult airways arise commonly. These data highlight the lack of availability of VL devices in these critical medical areas, which may be due to the relatively high cost of VL devices.¹¹

An effective low-cost VL device may help to increase availability of these devices in all areas where difficult airway management occurs, especially in lower income, resource-constrained countries where rescue airway devices are not readily available. Furthermore, reduction in cost may facilitate greater usage of VL, which could in turn improve patient safety and outcomes. In the context of the current global pandemic, video laryngoscopy has been advised in many airway guidelines.¹²⁻¹³ Access to a low-cost VL which does not require reprocessing may be of great value.

The Intubator V1-Indirect is a low-cost, disposable, hyperangulated VL developed in South Africa, which is not yet available on the market. The device consists of a single-piece, injection-moulded, polycarbonate blade-handle combination which houses a CMOS camera and reusable battery-powered electronics module. Insertion of the electronics module makes a video stream available via a 2.4 Ghz wireless access point. The user views the video stream using a purpose-built mobile phone application which presents the stream on the device display. The user is thus able to perform VL using a combination of the Intubator V1-Indirect and their smartphone. The device takes advantage of the processing and display capability of modern smartphones, including recording capability. The blade-handle-camera combination is disposable, eliminating the need for sterilisation and risks of spread of pathogens. The device is lightweight, robust, portable and compact. The quality of the video image produced with the endoscopic camera and mobile application is 720P, which exceeds the resolution of the comparator (CMAC).

(Figure 1 – Intubator V1 Indirect)

The aim of this study was to determine clinical equivalence of a low-cost video laryngoscope with a current standard device, using a simulated difficult airway model. Due to the hyperangulated shaped blade of the novel device, the Storz CMAC Pocket Monitor with Dörge blade (Karl Storz GmbH, Tuttlingen, Germany) was selected as an appropriate comparator. The Dörge blade is hyperangulated and designed for a difficult airway and anterior larynx.

(Figure 2 – Storz CMAC D-blade)

The null hypothesis is that a low cost video laryngoscope is not equivalent to the industry standard in simulated difficult airway. The primary outcome was mean difference in intubation times, with a secondary outcome of assessing laryngoscopic view obtained with each blade.

Methods

After approval by the University of Cape Town Human Research Ethics Committee (HREC REF 770/2018), a randomised, controlled, cross-over equivalence study was designed to compare the novel device to an established industry standard. Participant inclusion criteria were skilled intubators (>200 direct laryngoscopy intubations), with previous video laryngoscope experience.^{14 15} One hundred participants were recruited from Groote Schuur and Red Cross War Memorial Children's Hospitals with written informed consent between April and August 2019.

Participants received one-on-one demonstration of intubation with both the CMAC and low-cost device on an airway training mannikin (Laerdal, Stavanger, Norway) in a controlled environment. The manikin was immobilized with a cervical collar to simulate a difficult airway, as described in a previous study.¹⁶ Participants then had an opportunity to practice intubating the mannikin with a standard 7.0 mm internal diameter endotracheal tube loaded with a preformed malleable stylet with both devices. Participants could practice with the devices and were required to demonstrate at least three successful intubations with each device before they proceeded to perform their video-recorded intubations. The novel device had video streamed to a mobile phone (Samsung Galaxy S7 Android device), which participants could place at a convenient vantage point of their choice.

After practice attempts with each device, participants were randomised by blinded selection of a number from an envelope which determined the order of timed intubation attempts, with half randomised to use either the C-MAC or novel device first. Participants were asked to intubate the mannikin as swiftly as safely possible with each device and were not helped or guided during their intubation attempts. Intubation was then immediately performed with the alternate device.

Video recordings were uploaded to a notebook computer and imported to an editing suite (Sony Vegas 13C, Sony Creative Software Inc, Madison, U.S.A). Recordings were viewed frame-by-frame by two different observers. Time-point markers were placed at the start of intubation (deemed as the moment the device entered the oral cavity) and at the end of intubation (the moment the vocal cord marker on the ETT moved past the true cords). Laryngoscopic views were graded using a portion of the Fremantle score, as full, partial or no view of the glottic opening.¹⁷

Statistical methods:

A sample size of 97 participants was calculated using PASS software (PASS 15 Power Analysis and Sample Size Software 2017, NCSS, LLC. Kaysville, Utah, USA) for 90% power with an alpha error of 2.5%. Based on previous equivalence studies, *a priori* clinical equivalence limits were set at 10 seconds mean difference in intubation time.^{18 19} Intubation times were recorded on a Microsoft Excel spreadsheet (Microsoft Corporation, Redmond, Washington, USA), from where they were exported to MedCalc for statistical analysis (MedCalc® Statistical Software, Ostend, Belgium; <https://www.medcalc.org>; 2020). Interrater agreement was assessed using absolute agreement of intraclass correlation coefficient. The average of the times analysed by the two observers was used

in the final analysis. The mean difference in intubation time was compared using a two one-sided test (TOST).

Results

All participants were anaesthesia trainees and consultants with video laryngoscopic intubation experience. All successfully intubated the trachea during their timed attempt, with no oesophageal intubations with either device. The mean difference for time to intubation was 4.92 seconds (95% CI 2.34 - 7.49) as seen in Figure 2. This was within the 10 second *a priori* limits for clinical significance. The fastest time and slowest times with the C-MAC D blade were 5.1 and 34.2 seconds respectively, with an average time of 13.3 seconds. By comparison, the fastest and slowest time with the novel device were 5.2 and 99.1 seconds with an average time of 18.15 seconds (Table 1). Absolute interrater agreement in intubation between the two observers was high, with an intraclass correlation coefficient of 0.9945 (95% CI 0.9927 - 0.9958).

Views when using the CMAC were graded as full and partial in 42 and 58 cases respectively, and as full and partial in 41 and 59 cases when using the Intubator. There were no cases of no glottic view. (See Table 2).

Discussion

This study demonstrates equivalence of the low-cost video laryngoscope with an industry standard device in immobilised mannikins. Participants were able to learn how to use the new device within a brief period of time, and successfully intubate a simulated difficult airway. This study was in accordance with the ADEPT guidelines, which suggest that manikin evidence should exist before a new device is to be tested for equivalence in humans.²⁰

It is noteworthy that while the devices were clinically equivalent within our defined limits, the CMAC was nearly 5 seconds faster on average, with this result reaching statistical significance ($p = 0.0003$). Most of the participants in this study use the CMAC on a regular basis, as it is the video laryngoscope available in our institution. This may have led to some inherent performance bias in favour of the CMAC. However, despite a learning curve with new device, approximately one third (36/100) of participants were faster with the novel device than the CMAC.

Several other observations were made during the study in terms of the performance of the two devices. There was variability with the ease with which some participants learned how to use the novel device. Although all participants were allowed as much practice as desired and had to demonstrate at least three successful intubations prior to their recorded efforts, some participants may have rushed their training and practice attempts in order to return to their clinical duties. This may account for outliers who were significantly slower than average and may have needed more time to familiarize themselves with the new device.

The study naturally has several limitations. Participants were instructed to intubate the mannikin “as fast as possible and to the best of your ability” and the study relied heavily on the best effort during their timed intubation to give a true representation of the ability to use both devices. This may also have influenced the quality of the laryngoscopic views obtained, as participants were concentrating on swift intubation rather than obtaining the best possible view. The observers who analysed the recordings could not be blinded as to which device was in use. However, the excellent interrater agreement between observers suggests this had minimal effect. A few participants demonstrated

performance anxiety because they were being timed. We feel, however, that performing an emergency intubation also carries performance pressure, and thus this could still be an accurate reflection of their ability under stressful situations.

In accordance with previous studies in the literature, participants were experienced users. It is not known how the device might perform in the hands of inexperienced users. A subsequent study could assess how novice users might perform with two devices. The primary outcomes of this study was mean time to intubation, with laryngoscopic grading as a secondary outcome. Additional outcomes not assessed here which are often assessed in similar studies include ease of use, or airway/dental trauma.^{17 21} Although these metrics could provide additional information regarding the performance of the novel device, we selected equivalence in intubation time as the specific outcome relevant to securing an airway as swiftly as possible.

The Intubator-V1 was prototyped using 3D-printing, but blade used during the study was manufactured in the same fashion as the final production version, using injection moulding of polycarbonate, which is cheap but extremely durable. The same blade was used by all participants for multiple practice attempts and intubations throughout the study, and did not accrue any damage or experience technical issues during at least 400 uses. It therefore far exceeds robustness required of a single use device. It is noted that the environmental footprint of a disposable device could add to environmental cost. However, polycarbonate is fully recyclable, and lack of requirement for processing and sterilization decreases water and hazardous chemical use, reducing the environmental effects.

Difficult airway management is a skill required by expert anaesthesiologists and junior clinicians alike. Video laryngoscopy has proven to be a powerful modality in the field, and certainly has a place as a rescue device in unanticipated failed intubation with direct laryngoscopy.¹⁰ Furthermore, during the COVID-19 pandemic, various anaesthesia and airway societies have advocated for video laryngoscopy to help reduce healthcare worker exposure during intubation.^{12 22}

Studies have highlighted that VL is more readily available in operating theatre environments than in emergency departments and intensive care units, where difficult and unanticipated emergency intubation often takes place.¹¹ Difficult airway society guidelines now dictate that VL be available in all clinical areas where intubation takes place.² The major barrier to availability of these devices is financial constraints, especially in resource-constrained low- and middle-income countries, where studies show that perioperative outcomes are worse than the developed world.²³ The goal of the inventors of this low-cost device is to make video-laryngoscopy available to all medical practitioners, with a disposable, hyperangulated blade specifically designed for the difficult airway, or as a rescue device in the unanticipated difficult airway.

There is widespread acceptance that video laryngoscopy improves management of difficult airways, but the availability thereof is limited, particularly in resource-constrained environments. This discrepancy has been further highlighted during the COVID-19 pandemic, where most guidelines recommend the use of VL. A low-cost video laryngoscope could help bridge resource gap, and allow less well-funded areas and institutions to use this life saving equipment, and ultimately improve patient outcomes. This study has shown that such a device is clinically equivalent to the industry standard in a simulated difficult airway. This opens the way to further testing of this device *in vivo* to determine clinical utility and safety, and ultimately make the device available to Africa and the world.

Acknowledgements

The authors acknowledge the staff of Groote Schuur and Red Cross War Memorial Children's Hospitals, and the University of Cape Town Department of Anaesthesia and Perioperative Medicine for participating, and thank Karl Storz Endoscopy SA for the loan of a CMAC to perform this study. The study was supported by Departmental funds, including remuneration for a third-party statistician to confirm the results. We are grateful to Dr Dinell Behari for independent analysis of video footage to compile intubation times.

Competing interests

CdV and DE designed and produced the novel study device. RH is the director of the African Fellowship in Airway and Thoracic Anaesthesia, which is supported in part by an unrestricted educational grant by Karl Storz GbBH. CA has no competing interests to declare. All data and video files available at request of author.

References

1. Pieters BM, Eindhoven GB, Acott C, et al. Pioneers of laryngoscopy: indirect, direct and video laryngoscopy. *Anaesth Intensive Care* 2015;43 Suppl:4-11. doi: 10.1177/0310057X150430S103 [published Online First: 2015/07/01]
2. Frerk C, Mitchell VS, McNarry AF, et al. Difficult Airway Society 2015 guidelines for management of unanticipated difficult intubation in adults. *Br J Anaesth* 2015;115(6):827-48. doi: 10.1093/bja/aev371 [published Online First: 2015/11/12]
3. Apfelbaum JL, Hagberg CA, Caplan RA, et al. Practice guidelines for management of the difficult airway: an updated report by the American Society of Anesthesiologists Task Force on Management of the Difficult Airway. *Anesthesiology* 2013;118(2):251-70. doi: 10.1097/ALN.0b013e31827773b2 [published Online First: 2013/02/01]
4. Mushambi MC, Kinsella SM, Popat M, et al. Obstetric Anaesthetists' Association and Difficult Airway Society guidelines for the management of difficult and failed tracheal intubation in obstetrics. *Anaesthesia* 2015;70(11):1286-306. doi: 10.1111/anae.13260 [published Online First: 2015/10/10]
5. Chrimes N. The Vortex: a universal 'high-acuity implementation tool' for emergency airway management. *Br J Anaesth* 2016;117 Suppl 1:i20-i27. doi: 10.1093/bja/aew175 [published Online First: 2016/07/22]
6. Lewis SR, Butler AR, Parker J, et al. Videolaryngoscopy versus direct laryngoscopy for adult patients requiring tracheal intubation. *Cochrane Database Syst Rev* 2016;11:CD011136. doi: 10.1002/14651858.CD011136.pub2 [published Online First: 2016/11/16]
7. Michailidou M, O'Keeffe T, Mosier JM, et al. A comparison of video laryngoscopy to direct laryngoscopy for the emergency intubation of trauma patients. *World J Surg* 2015;39(3):782-8. doi: 10.1007/s00268-014-2845-z [published Online First: 2014/10/29]
8. De Jong A, Molinari N, Conseil M, et al. Video laryngoscopy versus direct laryngoscopy for orotracheal intubation in the intensive care unit: a systematic review and meta-analysis. *Intensive Care Med* 2014;40(5):629-39. doi: 10.1007/s00134-014-3236-5 [published Online First: 2014/02/22]
9. Sakles JC, Mosier J, Chiu S, et al. A comparison of the C-MAC video laryngoscope to the Macintosh direct laryngoscope for intubation in the emergency department. *Ann Emerg Med* 2012;60(6):739-48. doi: 10.1016/j.annemergmed.2012.03.031 [published Online First: 2012/05/09]
10. Kilicaslan A, Topal A, Tavlan A, et al. Effectiveness of the C-MAC video laryngoscope in the management of unexpected failed intubations. *Braz J Anesthesiol* 2014;64(1):62-5. doi: 10.1016/j.bjane.2013.03.001 [published Online First: 2014/02/26]
11. Cook TM, Kelly FE. A national survey of videolaryngoscopy in the United Kingdom. *Br J Anaesth* 2017;118(4):593-600. doi: 10.1093/bja/aex052 [published Online First: 2017/04/14]
12. Cook TM, El-Boghdady K, McGuire B, et al. Consensus guidelines for managing the airway in patients with COVID-19: Guidelines from the Difficult Airway Society, the Association of Anaesthetists the Intensive Care Society, the Faculty of Intensive Care Medicine and the Royal College of Anaesthetists. *Anaesthesia* 2020;75(6):785-99. doi: 10.1111/anae.15054
13. Matava CT, Kovatsis PG, Lee JK, et al. Pediatric Airway Management in COVID-19 Patients: Consensus Guidelines From the Society for Pediatric Anesthesia's Pediatric Difficult Intubation Collaborative and the Canadian Pediatric Anesthesia Society. *Anesth Analg* 2020;131(1):61-73. doi: 10.1213/ANE.0000000000004872 [published Online First: 2020/04/15]
14. Je S, Cho Y, Choi HJ, et al. An application of the learning curve-cumulative summation test to evaluate training for endotracheal intubation in emergency medicine. *Emerg Med J* 2015;32(4):291-4. doi: 10.1136/emermed-2013-202470 [published Online First: 2013/10/25]

15. Konrad C, Schupfer G, Wietlisbach M, et al. Learning manual skills in anesthesiology: Is there a recommended number of cases for anesthetic procedures? *Anesth Analg* 1998;86(3):635-9. doi: 10.1097/00000539-199803000-00037 [published Online First: 1998/03/12]
16. Kingma K, Hofmeyr R, Zeng IS, et al. Comparison of four methods of endotracheal tube passage in simulated airways: There is room for improved techniques. *Emerg Med Australas* 2017;29(6):650-57. doi: 10.1111/1742-6723.12874 [published Online First: 2017/11/08]
17. Swann AD, English JD, O'Loughlin EJ. The development and preliminary evaluation of a proposed new scoring system for videolaryngoscopy. *Anaesth Intensive Care* 2012;40(4):697-701. doi: 10.1177/0310057X1204000417 [published Online First: 2012/07/21]
18. Fiadjoe JE, Gurnaney H, Dalesio N, et al. A prospective randomized equivalence trial of the GlideScope Cobalt(R) video laryngoscope to traditional direct laryngoscopy in neonates and infants. *Anesthesiology* 2012;116(3):622-8. doi: 10.1097/ALN.0b013e318246ea4d
19. Jagannathan N, Hajduk J, Sohn L, et al. Randomized equivalence trial of the King Vision aBlade videolaryngoscope with the Miller direct laryngoscope for routine tracheal intubation in children <2 yr of age. *Br J Anaesth* 2017;118(6):932-37. doi: 10.1093/bja/aex073
20. Pandit JJ, Popat MT, Cook TM, et al. The Difficult Airway Society 'ADEPT' guidance on selecting airway devices: the basis of a strategy for equipment evaluation. *Anaesthesia* 2011;66(8):726-37. doi: 10.1111/j.1365-2044.2011.06787.x [published Online First: 2011/06/29]
21. Pieters BM, Wilbers NE, Huijzer M, et al. Comparison of seven videolaryngoscopes with the Macintosh laryngoscope in manikins by experienced and novice personnel. *Anaesthesia* 2016;71(5):556-64. doi: 10.1111/anae.13413 [published Online First: 2016/03/15]
22. De Jong A, Pardo E, Rolle A, et al. Airway management for COVID-19: a move towards universal videolaryngoscope? *Lancet Respir Med* 2020;8(6):555. doi: 10.1016/S2213-2600(20)30221-6
23. Biccadd BM, Madiba TE, Kluyts HL, et al. Perioperative patient outcomes in the African Surgical Outcomes Study: a 7-day prospective observational cohort study. *Lancet* 2018;391(10130):1589-98. doi: 10.1016/S0140-6736(18)30001-1

Figure 1. V1 Intubator Indirect - Novel low-cost hyperangulated disposable videolaryngoscope

Figure 2. Storz CMAC with Dörge's blade (Karl Storz GmbH, Tuttlingen, Germany)

Figure 3. Equivalence plot showing mean difference in intubation time. Red dotted lines show *a priori* clinical equivalence limit of 10 seconds.

CI = Confidence interval

Table 1. Intubation time between CMAC and Intubator V1 indirect video laryngoscope

	CMAC	Intubator V1	Difference
N	100	100	-
Minimum intubation time (seconds)	5,08	5,20	-
Maximum intubation time (seconds)	34,22	98,53	-
Mean intubation time (seconds)	13,31	18,23	4,92*
95% CI of intubation time	12,18 to 14,44	15,62 to 20, 84	2,34 to 7,49
Standard deviation of intubation time	5,70	13,14	12,97
Interquartile range of intubation time	9,25 to 16,61	10,65 to 20,37	-1,85 to 7,73

N: number, CI: confidence interval, SD: standard deviation, RSD: relative standard deviation, Normal Distr: Normal distribution). *Mean difference in intubation time $p = 0,0003$.

Table 2. Fremantle¹⁷ videolaryngoscopic score for view obtained with CMAC and Intubator.

Fremantle view grading	CMAC-D, n (%)	Intubator V-1, n (%)
Full	42 (42 %)	41 (41 %)
Partial	58 (58 %)	59 (59 %)
None	0 (0 %)	0 (0 %)
Total	100 (100 %)	100 (100 %)

Number of participants whom obtained a “Full”, “Partial” or “None” view as per Fremantle videolaryngoscopic grading.

Contents lists available at ScienceDirect

Trends in Anaesthesia and Critical Care

journal homepage: www.elsevier.com/locate/tacc Innovation in low-cost video-laryngoscopy: Intubator V1-Indirect compared with Storz C-MAC in a simulated difficult airway^{*}C.T. de Villiers^{*}, C. Alphonsus, D. Eave, R. Hofmeyr

Department of Anaesthesia and Perioperative Medicine, University of Cape Town, and Groote Schuur Hospital, Cape Town, South Africa

ARTICLE INFO

Article history:

Received 19 May 2021

Received in revised form

28 June 2021

Accepted 1 July 2021

Keywords:

Videolaryngoscopy

Video laryngoscope

Difficult airway

Low-cost

ABSTRACT

Background: Video laryngoscopy has directly impacted airway management, with numerous studies demonstrating its utility in clinical management of anatomically difficult airways. However, availability of video laryngoscopes in all clinical areas has been limited by the relatively high cost. We used smartphone technology, miniature cameras and three-dimensional printing to design and create an innovative low-cost hyperangulated video laryngoscope. This has the potential to make the technique more widely available. The aim of this study was to determine if time to intubation with the novel device was clinically equivalent to an existing gold-standard video laryngoscope (Storz CMAC with Dörge blade).

Methods: We conducted a randomised, controlled, cross-over equivalence study with 100 skilled practitioners who had previous video laryngoscopy experience. Participants received instruction on the new device, and adequate opportunity to practice. Intubations were then performed in a randomised order on a mannikin simulating a difficult airway. Video recordings of each intubation were analysed by two independent investigators to determine time to intubation. A mean difference in intubation time of less than 10 s was determined *a priori* to denote clinical equivalence.

Results: Mean difference in intubation time between the devices was 4.9 s, (two one-sided test 95%CI: 2.3–7.5 s). The innovative low-cost VL was thus clinically equivalent to the industry standard in a simulated difficult airway. Further testing *in vivo* in a clinical environment is needed.

Conclusion: The results of this study show that a low-cost disposable hyperangulated video laryngoscope is clinically equivalent to the industry standard in a simulated difficult airway. In the context of the current global pandemic, video laryngoscopy has been advised in nearly all airway guidelines. Access to a low-cost VL which does not require reprocessing may be of great value.

© 2021 Elsevier Ltd. All rights reserved.

1. Introduction

Traditionally, the most commonly used tool for manipulation of the airway and tracheal intubation under direct vision is a Macintosh blade laryngoscope [1]. The ability to safely perform this procedure requires practice and experience, and can prove challenging where there are anatomical or situational factors associated with a difficult airway [2–5].

In human trials, VLs have been shown to improve the view of the

larynx and glottis compared to direct laryngoscopy (DL).^{6 7} VLs have improved intubation first-pass success rates, decreased rates of oesophageal intubation, and increased overall intubation success for both expert and inexperienced providers, within as well as outside the operating theatre context [6–9]. Whilst DL intubation has been shown in some studies to be faster than VL, especially in experienced hands, VL may also be an important rescue technique for failed DL intubation. This is reflected in a study which showed the efficacy of the Storz CMAC video laryngoscope as a rescue device in the unanticipated difficult airway, where 86% first-pass success (FPS) was achieved after failed direct laryngoscopy [10]. The Difficult Airway Society (DAS) now recommends that video laryngoscopy should be immediately available whenever intubation is performed [11]. A United Kingdom national survey highlighted the fact that VLs are available in 93% of operating theatres, but only 54% of ICUs and 35% of emergency departments [11]. However, ICUs and emergency departments are clinical areas

^{*} This study was undertaken in partial completion of the MMed degree for the first author. Findings of this study have been presented in oral and poster format at the World Airway Management Meeting in Amsterdam, November 2019, and at the South African Society of Anaesthesiologists' annual congress in Pretoria, March 2020. The abstract for WAMM was published in *Trends in Anaesthesia and Critical Care* 2020(30):e127.

^{*} Corresponding author.

E-mail address: christiaandevilliers@gmail.com (C.T. de Villiers).

where unanticipated difficult airways arise commonly. These data highlight the lack of availability of VL devices in these critical medical areas, which may be due to the relatively high cost of VL devices [11].

An effective low-cost VL device may help to increase availability of these devices in all areas where difficult airway management occurs, especially in lower income, resource-constrained countries where rescue airway devices are not readily available. Furthermore, reduction in cost may facilitate greater usage of VL, which could in turn improve patient safety and outcomes. In the context of the current global pandemic, video laryngoscopy has been advised in many airway guidelines.[12 13] Access to a low-cost VL which does not require reprocessing may be of great value.

The Intubator V1-Indirect is a low-cost, disposable, hyper-angulated VL developed in South Africa, which is not yet available on the market. The device consists of a single-piece, injection-moulded, polycarbonate blade-handle combination which houses a CMOS camera and reusable battery-powered electronics module (see Fig. 1). Insertion of the electronics module makes a video stream available via a 2.4 Ghz wireless access point. The user views the video stream using a purpose-built mobile phone application which presents the stream on the device display. The user is thus able to perform VL using a combination of the Intubator V1-Indirect and their smartphone. The device takes advantage of the processing and display capability of modern smartphones, including recording capability. The blade-handle-camera combination is disposable, eliminating the need for sterilization and risks of spread of pathogens. The device is lightweight, robust, portable and compact. The quality of the video image produced with the endoscopic camera and mobile application is 720 P, which exceeds the resolution of the comparator (CMAC).

The aim of this study was to determine clinical equivalence of a low-cost video laryngoscope with a current standard device, using a simulated difficult airway model. Due to the hyperangulated

shaped blade of the novel device, the Storz CMAC Pocket Monitor with Dörge blade (Karl Storz GmbH, Tuttlingen, Germany) was selected as an appropriate comparator. The Dörge blade is hyper-angulated and designed for a difficult airway and anterior larynx (see Fig. 2).

The null hypothesis is that a low cost video laryngoscope is not equivalent to the industry standard in simulated difficult airway. The primary outcome was mean difference in intubation times, with a secondary outcome of assessing laryngoscopic view obtained with each blade.

2. Methods

After approval by the University of Cape Town Human Research Ethics Committee (HREC REF 770/2018), a randomised, controlled, cross-over equivalence study was designed to compare the novel device to an established industry standard. Participant inclusion criteria were skilled intubators (>200 direct laryngoscopy intubations), with previous video laryngoscope experience.[14 15] One hundred participants were recruited from Groote Schuur and Red Cross War Memorial Children's Hospitals with written informed consent between April and August 2019.

Participants received one-on-one demonstration of intubation with both the CMAC and low-cost device on an airway training mannikin (Laerdal, Stavanger, Norway) in a controlled environment. The manikin was immobilized with a cervical collar to simulate a difficult airway, as described in a previous study [16]. Participants then had an opportunity to practice intubating the mannikin with a standard 7.0 mm internal diameter endotracheal tube loaded with a preformed malleable stylet with both devices. Participants could practice with the devices and were required to demonstrate at least three successful intubations with each device before they proceeded to perform their video-recorded intubations. The novel device had video streamed to a mobile phone (Samsung Galaxy S7 Android device), which participants could place at a convenient vantage point of their choice.

After practice attempts with each device, participants were randomised by blinded selection of a number from an envelope

Fig. 1. V1 Intubator Indirect - Novel low-cost hyperangulated disposable videolaryngoscope.

Fig. 2. Storz CMAC with Dörge blade (Karl Storz GmbH, Tuttlingen, Germany).

which determined the order of timed intubation attempts, with half randomised to use either the C-MAC or novel device first. Participants were asked to intubate the mannikin as swiftly as safely possible with each device and were not helped or guided during their intubation attempts. Intubation was then immediately performed with the alternate device.

Video recordings were uploaded to a notebook computer and imported to an editing suite (Sony Vegas 13C, Sony Creative Software Inc, Madison, U.S.A). Recordings were viewed frame-by-frame by two different observers. Time-point markers were placed at the start of intubation (deemed as the moment the device entered the oral cavity) and at the end of intubation (the moment the vocal cord marker on the ETT moved passed the true cords). Laryngoscopic views were graded using a portion of the Fremantle score, as full, partial or no view of the glottic opening [17].

2.1. Statistical methods

A sample size of 97 participants was calculated using PASS software (PASS 15 Power Analysis and Sample Size Software 2017, NCSS, LLC. Kaysville, Utah, USA) for 90% power with an alpha error of 2.5%. Based on previous equivalence studies, *a priori* clinical equivalence limits were set at 10 s mean difference in intubation time.[18–19] Intubation times were recorded on a Microsoft Excel spreadsheet (Microsoft Corporation, Redmond, Washington, USA), from where they were exported to MedCalc for statistical analysis (MedCalc® Statistical Software, Ostend, Belgium; <https://www.medcalc.org>; 2020). Interrater agreement was assessed using absolute agreement of intraclass correlation coefficient. The average of the times analysed by the two observers was used in the final analysis. The mean difference in intubation time was compared using a two one-sided test (TOST).

3. Results

All participants were anaesthesia trainees and consultants with video laryngoscopic intubation experience. All successfully intubated the trachea during their timed attempt, with no oesophageal intubations with either device. The mean difference for time to intubation was 4.9 s (95% CI 2.3–7.5) as seen in Fig. 3. This was within the 10 s *a priori* limits for clinical significance. The fastest time and slowest times with the C-MAC D blade were 5.1 and 34.2 s respectively, with an average time of 13.3 s. By comparison, the fastest and slowest time with the novel device were 5.2 and 99.1 s with an average time of 18.2 s (Table 1). Absolute interrater agreement in intubation between the two observers was high, with an intraclass correlation coefficient of 0.995 (95% CI 0.993–0.996).

Views when using the CMAC were graded as full and partial in 42 and 58 cases respectively, and as full and partial in 41 and 59 cases when using the Intubator. There were no cases of no glottic view. (See Table 2).

4. Discussion

This study demonstrates equivalence of the low-cost video laryngoscope with an industry standard device in immobilized mannikins. Participants were able to learn how to use the new device within a brief period of time, and successfully intubate a simulated difficult airway. This study was in accordance with the ADEPT guidelines, which suggest that manikin evidence should exist before a new device is to be tested for equivalence in humans [20].

It is noteworthy that while the devices were clinically equivalent within our defined limits, the CMAC was nearly 5 s faster on average, with this result reaching statistical significance

($p = 0.0003$). Most of the participants in this study use the CMAC on a regular basis, as it is the video laryngoscope available in our institution. This may have led to some inherent performance bias in favour of the CMAC. However, despite a learning curve with new device, approximately one third (36/100) of participants were faster with the novel device than the CMAC.

Several other observations were made during the study in terms of the performance of the two devices. There was variability with the ease with which some participants learned how to use the novel device. Although all participants were allowed as much practice as desired and had to demonstrate at least three successful intubations prior to their recorded efforts, some participants may have rushed their training and practice attempts in order to return to their clinical duties. This may account for outliers who were significantly slower than average and may have needed more time to familiarize themselves with the new device.

The study naturally has several limitations. Participants were instructed to intubate the mannikin “as fast as possible and to the best of your ability” and the study relied heavily on the best effort during their timed intubation to give a true representation of the ability to use both devices. This may also have influenced the quality of the laryngoscopic views obtained, as participants were concentrating on swift intubation rather than obtaining the best possible view. The observers who analysed the recordings could not be blinded as to which device was in use. However, the excellent interrater agreement between observers suggests this had minimal effect. A few participants demonstrated performance anxiety because they were being timed. We feel, however, that performing an emergency intubation also carries performance pressure, and thus this could still be an accurate reflection of their ability under stressful situations.

In accordance with previous studies in the literature, participants were experienced users. It is not known how the device might perform in the hands of inexperienced users. A subsequent study could assess how novice users might perform with two devices. The primary outcomes of this study was mean time to intubation, with laryngoscopic grading as a secondary outcome. Additional outcomes not assessed here which are often assessed in similar studies include ease of use, or airway/dental trauma.[17–21] Although these metrics could provide additional information regarding the performance of the novel device, we selected equivalence in intubation time as the specific outcome relevant to securing an airway as swiftly as possible.

The Intubator-V1 was prototyped using 3D-printing, but blade used during the study was manufactured in the same fashion as the final production version, using injection molding of polycarbonate, which is cheap but extremely durable. The same blade was used by all participants for multiple practice attempts and intubations throughout the study, and did not accrue any damage or experience technical issues during at least 400 uses. It therefore far exceeds robustness required of a single use device. It is noted that the environmental footprint of a disposable device could add to environmental cost. However, polycarbonate is fully recyclable, and lack of requirement for processing and sterilization decreases water and hazardous chemical use, reducing the environmental effects.

Difficult airway management is a skill required by expert anaesthesiologists and junior clinicians alike. Video laryngoscopy has proven to be a powerful modality in the field, and certainly has a place as a rescue device in unanticipated failed intubation with direct laryngoscopy [10]. Furthermore, during the COVID-19 pandemic, various anaesthesia and airway societies have advocated for video laryngoscopy to help reduce healthcare worker exposure during intubation.[12–22]

Studies have highlighted that VL is more readily available in operating theatre environments than in emergency departments

Fig. 3. Equivalence plot showing mean difference in intubation time. Red dotted lines show *a priori* clinical equivalence limit of 10 s. (For interpretation of the references to colour in this figure legend, the reader is referred to the Web version of this article.)

Table 1
Intubation time between CMAC and Intubator V1 indirect video laryngoscope.

	CMAC	Intubator V1	Difference
N	100	100	–
Minimum – maximum intubation time (seconds)	5,1–34,2	5,2–99,1	–
intubation time (seconds) mean (SD)	13,3 (5,7)	18,2 (13,1)	4,9 ^a
95% CI of intubation time	12,2 to 14,4	15,6 to 20, 8	2,3 to 7,5

N: number, CI: confidence interval, SD: standard deviation.

^a Mean difference in intubation time $p = 0,0003$.

Table 2
Fremantle [17] videolaryngoscopic score for view obtained with CMAC and Intubator.

Freemantle view grading	CMAC-D, n (%)	Intubator V-1, n (%)
Full	42 (42%)	41 (41%)
Partial	58 (58%)	59 (59%)
None	0 (0%)	0 (0%)
Total	100 (100%)	100 (100%)

Number of participants whom obtained a “Full”, “Partial” or “None” view as per Fremantle videolaryngoscopic grading.

and intensive care units, where difficult and unanticipated emergency intubation often takes place [11]. Difficult airway society guidelines now dictate that VL be available in all clinical areas where intubation takes place [2]. The major barrier to availability of these devices is financial constraints, especially in resource-constrained low- and middle-income countries, where studies show that perioperative outcomes are worse than the developed world [23]. The goal of the inventors of this low-cost device is to make video-laryngoscopy available to all medical practitioners, with a disposable, hyperangulated blade specifically designed for the difficult airway, or as a rescue device in the unanticipated difficult airway.

There is widespread acceptance that video laryngoscopy improves management of difficult airways, but the availability thereof is limited, particularly in resource-constrained environments. This discrepancy has been further highlighted during the COVID-19 pandemic, where most guidelines recommend the use of VL. A low-cost video laryngoscope could help bridge resource gap, and allow less well-funded areas and institutions to use this life saving equipment, and ultimately improve patient outcomes. This study has shown that such a device is clinically equivalent to the industry standard in a simulated difficult airway. This opens the way to further testing of this device *in vivo* to determine clinical utility and safety, and ultimately make the device available to Africa and the world.

Author statement

Christiaan de Villiers: Conception or design of the work, data collection, data analysis and interpretation, drafting the article. Christella Alphonsus: analysis and interpretation, drafting the article. Dylan Eave: Conception or design of the work, drafting the article. Ross Hofmeyr: Conception or design of the work, data analysis and interpretation critical revision of the article.

Declaration of competing interest

CdV and DE designed and produced the novel study device. RH is the director of the African Fellowship in Airway and Thoracic Anaesthesia, which is supported in part by an unrestricted educational grant by Karl Storz GbBH. CA has no competing interests to declare. All data and video files available at request of author.

Acknowledgements

The authors acknowledge the staff of Groote Schuur and Red Cross War Memorial Children's Hospitals, and the University of Cape Town Department of Anaesthesia and Perioperative Medicine for participating, and thank Karl Storz Endoscopy SA for the loan of a CMAC to perform this study. The study was supported by Departmental funds, including remuneration for a third-party statistician to confirm the results. We are grateful to Dr Dinell Behari for independent analysis of video footage to compile intubation times.

References

- [1] B.M. Pieters, G.B. Eindhoven, C. Acott, et al., *Pioneers of laryngoscopy: indirect, direct and video laryngoscopy*, *Anaesth. Intensive Care* 43 (Suppl) (2015) 4–11, <https://doi.org/10.1177/0310057X1504305103> [published Online First: 2015/07/01].
- [2] C. Frerk, V.S. Mitchell, A.F. McNarry, et al., *Difficult Airway Society 2015 guidelines for management of unanticipated difficult intubation in adults*, *Br. J. Anaesth.* 115 (6) (2015) 827–848, <https://doi.org/10.1093/bja/aev371> [published Online First: 2015/11/12].
- [3] J.L. Apfelbaum, C.A. Hagberg, R.A. Caplan, et al., *Practice guidelines for management of the difficult airway: an updated report by the American society of anesthesiologists task force on management of the difficult airway*, *Anesthesiology* 118 (2) (2013) 251–270, <https://doi.org/10.1097/ALN.0-b013e31827773b2> [published Online First: 2013/02/01].
- [4] M.C. Mushambi, S.M. Kinsella, M. Papat, et al., *Obstetric Anaesthetists' Association and Difficult Airway Society guidelines for the management of difficult and failed tracheal intubation in obstetrics*, *Anaesthesia* 70 (11) (2015) 1286–1306, <https://doi.org/10.1111/anae.13260> [published Online First: 2015/10/10].
- [5] N. Chrimes, *The Vortex: a universal 'high-acuity implementation tool' for emergency airway management*, *Br. J. Anaesth.* 117 (1) (2016) i20–i27, <https://doi.org/10.1093/bja/aew175> [published Online First: 2016/07/22].
- [6] S.R. Lewis, A.R. Butler, J. Parker, et al., *Videolaryngoscopy versus direct laryngoscopy for adult patients requiring tracheal intubation*, *Cochrane Database Syst. Rev.* 11 (2016) CD011136, <https://doi.org/10.1002/14651858.CD011136.pub2> [published Online First: 2016/11/16].
- [7] M. Michailidou, T. O'Keeffe, J.M. Mosier, et al., *A comparison of video laryngoscopy to direct laryngoscopy for the emergency intubation of trauma patients*, *World J. Surg.* 39 (3) (2015) 782–788, <https://doi.org/10.1007/s00268-014-2845-z> [published Online First: 2014/10/29].
- [8] A. De Jong, N. Molinari, M. Conseil, et al., *Video laryngoscopy versus direct laryngoscopy for orotracheal intubation in the intensive care unit: a systematic review and meta-analysis*, *Intensive Care Med.* 40 (5) (2014) 629–639, <https://doi.org/10.1007/s00134-014-3236-5> [published Online First: 2014/02/22].
- [9] J.C. Sakles, J. Mosier, S. Chiu, et al., *A comparison of the C-MAC video laryngoscope to the Macintosh direct laryngoscope for intubation in the emergency department*, *Ann. Emerg. Med.* 60 (6) (2012) 739–748, <https://doi.org/10.1016/j.annemergmed.2012.03.031> [published Online First: 2012/05/09].
- [10] A. Kilicaslan, A. Topal, A. Tavlan, et al., *Effectiveness of the C-MAC video laryngoscope in the management of unexpected failed intubations*, *Braz J Anesthesiol* 64 (1) (2014) 62–65, <https://doi.org/10.1016/j.bjane.2013.03.001> [published Online First: 2014/02/26].
- [11] T.M. Cook, F.E. Kelly, *A national survey of videolaryngoscopy in the United Kingdom*, *Br. J. Anaesth.* 118 (4) (2017) 593–600, <https://doi.org/10.1093/bja/aex052> [published Online First: 2017/04/14].
- [12] T.M. Cook, K. El-Boghdady, B. McGuire, et al., *Consensus guidelines for managing the airway in patients with COVID-19: guidelines from the difficult airway society, the association of anaesthetists the intensive care society, the faculty of intensive care medicine and the royal college of anaesthetists*, *Anaesthesia* 75 (6) (2020) 785–799, <https://doi.org/10.1111/anae.15054>.
- [13] C.T. Matava, P.G. Kovatsis, J.K. Lee, et al., *Pediatric airway management in COVID-19 patients: consensus guidelines from the society for pediatric anaesthesia's pediatric difficult intubation collaborative and the Canadian pediatric anaesthesia society*, *Anesth. Analg.* 131 (1) (2020) 61–73, <https://doi.org/10.1213/ANE.00000000000004872> [published Online First: 2020/04/15].
- [14] S. Je, Y. Cho, H.J. Choi, et al., *An application of the learning curve-cumulative summation test to evaluate training for endotracheal intubation in emergency medicine*, *Emerg. Med. J.* 32 (4) (2015) 291–294, <https://doi.org/10.1136/emered-2013-202470> [published Online First: 2013/10/25].
- [15] C. Konrad, G. Schupfer, M. Wietlisbach, et al., *Learning manual skills in anesthesiology: is there a recommended number of cases for anesthetic procedures?* *Anesth. Analg.* 86 (3) (1998) 635–639, <https://doi.org/10.1097/0000539-199803000-00037> [published Online First: 1998/03/12].
- [16] K. Kingma, R. Hofmeyr, I.S. Zeng, et al., *Comparison of four methods of endotracheal tube passage in simulated airways: there is room for improved techniques*, *Emerg. Med. Australasia (EMA)* 29 (6) (2017) 650–657, <https://doi.org/10.1111/1742-6723.12874> [published Online First: 2017/11/08].
- [17] A.D. Swann, J.D. English, E.J. O'Loughlin, *The development and preliminary evaluation of a proposed new scoring system for videolaryngoscopy*, *Anaesth. Intensive Care* 40 (4) (2012) 697–701, <https://doi.org/10.1177/0310057X1204000417> [published Online First: 2012/07/21].
- [18] J.E. Fiadjo, H. Gurnaney, N. Dalesio, et al., *A prospective randomized equivalence trial of the GlideScope Cobalt(R) video laryngoscope to traditional direct laryngoscopy in neonates and infants*, *Anesthesiology* 116 (3) (2012) 622–628, <https://doi.org/10.1097/ALN.0b013e318246ea4d>.
- [19] N. Jagannathan, J. Hajduk, L. Sohn, et al., *Randomized equivalence trial of the King Vision aBlade videolaryngoscope with the Miller direct laryngoscope for routine tracheal intubation in children <2 yr of age*, *Br. J. Anaesth.* 118 (6) (2017) 932–937, <https://doi.org/10.1093/bja/aex073>.
- [20] J.J. Pandit, M.T. Papat, T.M. Cook, et al., *The Difficult Airway Society 'ADEPT' guidance on selecting airway devices: the basis of a strategy for equipment evaluation*, *Anaesthesia* 66 (8) (2011) 726–737, <https://doi.org/10.1111/j.1365-2044.2011.06787.x> [published Online First: 2011/06/29].
- [21] B.M. Pieters, N.E. Wilbers, M. Huijzer, et al., *Comparison of seven videolaryngoscopes with the Macintosh laryngoscope in manikins by experienced and novice personnel*, *Anaesthesia* 71 (5) (2016) 556–564, <https://doi.org/10.1111/anae.13413> [published Online First: 2016/03/15].
- [22] A. De Jong, E. Pardo, A. Rolle, et al., *Airway management for COVID-19: a move towards universal videolaryngoscopy?* *Lancet Respir Med* 8 (6) (2020) 555, [https://doi.org/10.1016/S2213-2600\(20\)30221-6](https://doi.org/10.1016/S2213-2600(20)30221-6).
- [23] B.M. Biccard, T.E. Madiba, H.L. Kluyts, et al., *Perioperative patient outcomes in the African Surgical Outcomes Study: a 7-day prospective observational cohort study*, *Lancet* 391 (10130) (2018) 1589–1598, [https://doi.org/10.1016/S0140-6736\(18\)30001-1](https://doi.org/10.1016/S0140-6736(18)30001-1).

List of appendices

Appendix 1: TACC Publication acceptance

Appendix 2: TACC Reviewer Comments

Appendix 3: Abstract Publication TACC

Appendix 4: Human Research Ethics Committee approvals

Appendix 5: Mannikin study consent form

Appendix 6: Participant data capture form

Appendix 7: TACC Guide to authors

Appendix 1: TACC Acceptance for Publication

Christiaan de Villiers <christiaandevilliers@gmail.com>

Decision on submission to Trends in Anaesthesia and Critical Care

1 message

Trends in Anaesthesia and Critical Care <em@editorialmanager.com>

Thu, Jul 1, 2021 at 6:06 PM

Reply-To: Trends in Anaesthesia and Critical Care <tacc@elsevier.com>

To: Christiaan de Villiers <christiaandevilliers@gmail.com>

CC: zoltanszucsmd@gmail.com

Manuscript Number: TACC-D-21-00083R1

Innovation in low-cost video laryngoscopy: Intubator V1-Indirect compared with Storz C-MAC in a simulated difficult airway

Dear Dr de Villiers,

Thank you for submitting your manuscript to Trends in Anaesthesia and Critical Care.

I am pleased to inform you that your manuscript has been accepted for publication. Your accepted manuscript will now be transferred to our production department. We will create a proof which you will be asked to check, and you will also be asked to complete a number of online forms required for publication. If we need additional information from you during the production process, we will contact you directly.

We appreciate and value your contribution to Trends in Anaesthesia and Critical Care. We regularly invite authors of recently published manuscript to participate in the peer review process. If you were not already part of the journal's reviewer pool, you have now been added to it. We look forward to your continued participation in our journal, and we hope you will consider us again for future submissions.

Kind regards,
Robert Greif, MD, MME
Editor-in-Chief

Trends in Anaesthesia and Critical Care

Editor comments:

Dear author,

The changes requested by the reviewers have been expanded and edited in a correct manner. Excellent, interesting and useful work. Minor discrepancies remained in the numbers between the table and the results section. After editorial correction, I think the manuscript is already suitable for publication.

Appendix 2: TACC Reviewer Comments

Christiaan de Villiers <christiaandevilliers@gmail.com>

Decision on submission to Trends in Anaesthesia and Critical Care

6 messages

Trends in Anaesthesia and Critical Care <em@editorialmanager.com>

Fri, Jun 11, 2021 at 1:08 PM

Reply-To: Trends in Anaesthesia and Critical Care <tacc@elsevier.com>

To: Christiaan de Villiers <christiaandevilliers@gmail.com>

Manuscript Number: TACC-D-21-00083

Innovation in low-cost video laryngoscopy: Intubator V1-Indirect compared with Storz C-MAC in a simulated difficult airway

Dear Dr de Villiers,

Thank you for submitting your manuscript to Trends in Anaesthesia and Critical Care.

I have completed my evaluation of your manuscript. The reviewers recommend reconsideration of your manuscript following minor revision and modification. I invite you to resubmit your manuscript after addressing the comments below. Please resubmit your revised manuscript by Jul 11, 2021.

When revising your manuscript, please consider all issues mentioned in the reviewers' comments carefully: please outline every change made in response to their comments and provide suitable rebuttals for any comments not addressed. Please note that your revised submission may need to be re-reviewed.

To submit your revised manuscript, please log in as an author at <https://www.editorialmanager.com/tacc/>, and navigate to the "Submissions Needing Revision" folder under the Author Main Menu.

Trends in Anaesthesia and Critical Care values your contribution and I look forward to receiving your revised manuscript.

Kind regards,

Robert Greif, MD, MME

Editor-in-Chief

Trends in Anaesthesia and Critical Care

Editor and Reviewer comments:

Reviewer #1: Innovation in low-cost video laryngoscopy: Intubator V1-Indirect compared with Storz C-MAC in a simulated difficult airway

Overall a well conducted manikin study of comparing intubator 1 with C-MAC.

1. Introduction can be further improved. Second paragraph in the introduction is not necessary and can be removed.
2. The background on benefits of video laryngoscopy can be more concise. One paragraph with key references for VLs improve the success rate, shorter intubation time...etc should be fine.
3. I suggest that the paragraph describing the Intubator V1-direct from methods section to be moved to the introduction
4. I suggest to revise the introduction as below
First paragraph: the existing paragraph is fine, second paragraph: Concise benefits of VL, third paragraph:

Intubator V1, fourth Paragraph: The reason for choosing CMAC and aim of the study

Results:

1. Please provide the grade and clinical experience of the participants. Did all participants intubated the trachea in first attempt?
2. The second paragraph, the first sentence should be moved to Methods section, How you scored the laryngoscopy view should be under methods, provide reference to Fremantle score.
3. Table1. The 4th column second row, under difference 100 (should it be 0 or leave the cell blank). Also provide p value where applicable
4. Table 2. The title is incomplete Please provide more details to the title, not sure what those numbers in the table are. Also provide p value.

Discussion: Fifth paragraph, end of 4th line a typo, delete repeated word not

Reviewer #2: The current version of the manuscript is not suitable for publication and I urge the authors to basically change the entire manuscript, as there are major flaws:

Introduction:

The introduction needs to be re-written. Only some selective publications have been cited, although there are much more recent and larger trials in the field. Several strong statements, like "for both expert and inexperienced providers" are not supported by adequate references. Authors have at least mention, that there are also some controversial publications, especially dealing with experienced providers and do not indicate any superiority of videolaryngoscopy over direct laryngoscopy. Another example in the fifth paragraph "has been advised in nearly all airway guidelines". Please provide adequate references to support this statement.

References are used randomly. For example first sentence of the introduction. Although most readers might agree, the reference should not focus on pediatric intubations.

The introduction should inform readers about current knowledge, but should generally not focus on a few studies only. Instead of reporting too much details about a few studies, authors should provide a summary and indicate, why this study is needed to gain more knowledge/evidence.

Although the aim of this project is listed at the end of the introduction, neither the hypothesis nor the definitions of the outcomes are provided throughout the manuscript. This is quite unacceptable at this stage of the submission process.

Methods:

This section should be separated in at least 4 parts: Study setting/ study population, Study protocol, measurement, statistical methods. Please revise the current form.

Further comments:

Time when the study was performed is missing.

Study project was not a priori registered in clinicaltrials.gov.

I don't practice in Europe and I therefore have no idea about the Doerges blade. Please provide more information's about this blade.

Informations about the blade should be reported in the introduction, NOT in the methods.

Please provide further informations about the randomization, When? By whom?

Outcomes including exact definitions are still unclear. I assume that time to intubation was chosen as the primary outcome, which is regretful. Although time to intubation is commonly used in airway trials, the clinical relevance and importance is questionable. Most of clinicians would agree, that a difference of even 10 seconds might be highly statistical significant, but clinically absolutely unimportant. Safety, airway injuries, number of intubation attempts, intubation failure, and unrecognized esophageal intubations are much more important clinical outcomes.

Statistical analysis is basically missing at all.

Results:

It is still unclear to this reviewer, which outcomes have been defined a priori.

Authors should not provide any discussion like "it is thus..." in this section ☐ please move this to the discussion section.

Laryngoscopic view is definitely not fair, Anesthesia providers were asked to intubate as soon as possible, but were not asked to provide the best visualization possible. Therefore, providers might not have been focusing on obtaining better visualization, although it would have been possible by maybe even minor improvements.

Discussion:

This needs to be rewritten. Please focus on the current study and discuss the new findings in the context of currently known evidence. Authors do not have to defend this study, or why it is so important to do this study (□ this should be clear after the introduction).

Figure 1. I would suggest to add another picture of the device from the side to indicate the exact curvature of the blade. Maybe also add the comparison blade, so that reader have an idea, if these blades are comparable.

Figure 2. Although technically correct, this figure does not add anything. I would suggest to delete.

Table 1. was performed in a sloppy style. There are several empty rows at the bottom.

Table 2. Why did the authors choose to report the fremantle score? POGO or even modified C+L would be alternatives

I am just wondering, but is this device protected by any patent?

TACC-D-21-00083 – Authors’ response to reviewers

Thank you for your review, and the comments below. We have responded in **red/bold** to each comment, and highlighted the changes in the accompanying revised manuscript.

Reviewer #1: Innovation in low-cost video laryngoscopy: Intubator V1-Indirect compared with Storz C-MAC in a simulated difficult airway

Overall a well conducted manikin study of comparing intubator 1 with C-MAC.

>> Thank you.

1. Introduction can be further improved. Second paragraph in the introduction is not necessary and can be removed.

>> Introduction amended as per suggestion.

>> 2nd paragraph removed.

2. The background on benefits of video laryngoscopy can be more concise. One paragraph with key references for VLs improve the success rate, shorter intubation time...etc should be fine.

>>Background section revised.

3. I suggest that the paragraph describing the Intubator V1-direct from methods section to be moved to the introduction

>>Moved to introduction section as suggested

4. I suggest to revise the introduction as below

First paragraph: the existing paragraph is fine,

second paragraph: Concise benefits of VL,

third paragraph: Intubator V1,

fourth Paragraph: The reason for choosing CMAC and aim of the study

>> 1st paragraph unchanged as suggested

>> 2nd paragraph as suggested

>> 3rd paragraph... Intubator V1 moved from Methods to introduction as suggested

>> 4th paragraph...Aim of study

Results:

1. Please provide the grade and clinical experience of the participants.

>> Participants clinical experience updated

Did all participants intubated the trachea in first attempt?

>> Yes, this has been added.

2. The second paragraph, the first sentence should be moved to Methods section. How you scored the laryngoscopy view should be under methods, provide reference to Fremantle score.

>> Thank you; this has been amended and added.

3. Table1. The 4th column second row, under difference 100 (should it be 0 or leave the cell blank). Also provide p value where applicable

>> Thank you – the table has been altered to indicate where the “difference” is not applicable, and a p-value for the mean difference in intubation time has been added to the comments.

4. Table 2. The title is incomplete
Please provide more details to the title,

>> Amended

not sure what those numbers in the table are.

>> Amended labels to improve clarity

Also provide p value.

>> We have not performed any statistical analysis on this data, as the study was not powered for this outcome, and it is obvious to the reader that the numbers are nearly identical.

Discussion: Fifth paragraph, end of 4th line a typo, delete repeated word not

>> Amended

Reviewer #2: The current version of the manuscript is not suitable for publication and I urge the authors to basically change the entire manuscript, as there are major flaws:

Introduction:

The introduction needs to be re-written.

Only some selective publications have been cited, although there are much more recent and larger trials in the field.

>> We appreciate the reviewer's candour. We have taken pains to re-write the introduction and move sections in line with the other reviewer's comments, as well as reviewing and changing the referencing to be more specific.

Several strong statements, like "for both expert and inexperienced providers" are not supported by adequate references.

>> We have provided 4 references in support of this statement.

Authors have at least mention, that there are also some controversial publications, especially dealing with experienced providers and do not indicate any superiority of videolaryngoscopy over direct laryngoscopy.

>> We have added a comment at the start of the third paragraph to indicate this.

Another example in the fifth paragraph "has been advised in nearly all airway guidelines". Please provide adequate references to support this statement.

>> X2 Reference added

References are used randomly. For example first sentence of the introduction. Although most readers might agree, the reference should not focus on pediatric intubations.

>> Reference changed

The introduction should inform readers about current knowledge, but should generally not focus on a few studies only. Instead of reporting too much details about a few studies, authors should provide a summary and indicate, why this study is needed to gain more knowledge/evidence.

>> Introduction more concise and summarised

Although the aim of this project is listed at the end of the introduction, neither the hypothesis

>> null hypothesis added

nor the definitions of the outcomes are provided throughout the manuscript. This is quite unacceptable at this stage of the submission process.

>> The primary outcome is stated at the end of the introduction. We have added the secondary outcome to this paragraph. The specifics of measuring these outcomes are listed in the Methods section.

Methods:

This section should be separated in at least 4 parts:

Study setting/ study population,
Study protocol,
measurement,
statistical methods.
Please revise the current form.

>> Thank you – this is not specified in the instructions to authors, but we have revised as per the reviewer's suggestion

Further comments:

Time when the study was performed is missing.

>> Thank you, this has been added

Study project was not a priori registered in clinicaltrials.gov.

>> This study was not a clinical trial; the intervention was performed solely on manikins.

I don't practice in Europe and I therefore have no idea about the Doerges blade. Please provide more information's about this blade.

>> It is unfortunate that the reviewer is unfamiliar with the CMAC system and Dörge blade, as these are used widely around the world. It is worth noting that the authors, too, do not practice in Europe. We have included further information for the reader.

Informations about the blade should be reported in the introduction, NOT in the methods.

>> This has been changed

Please provide further informations about the randomization. When? By whom?

>> This has been added

Outcomes including exact definitions are still unclear. I assume that time to intubation was chosen as the primary outcome, which is regretful. Although time to intubation is commonly used in airway trials, the clinical relevance and importance is questionable. Most of clinicians would agree, that a difference of even 10 seconds might be highly statistical significant, but clinically absolutely unimportant. Safety, airway injuries, number of intubation attempts, intubation failure, and unrecognized esophageal intubations are much more important clinical outcomes.

>> We note the reviewer's comments. In terms of outcome definitions, we refer the reviewer to the methods section where we describe the frame-by-frame review of the video recordings to obtain precise intubation times. As the reviewer indicates, time to intubation is widely used in intubation studies, as it is a continuous variable and useful for statistical analysis. As all participants successfully intubated the trachea on the first

attempt, neither intubation success rate nor number of attempts is a useful or comparable outcome. There were no oesophageal intubations with either blades. Assessing "safety" and airway injuries is beyond the scope of this study, and not readily achievable in a manikin model.

Statistical analysis is basically missing at all.

>> We are unsure what details the reviewer is missing. The sample size calculation and methods are described in the methods section, as are the statistical methods for assessing interrater agreement and mean difference in intubation time (two one-sided test).

Results:

It is still unclear to this reviewer, which outcomes have been defined a priori.

Authors should not provide any discussion like "it is thus..." in this section ◊ please move this to the discussion section.

>> As stated in the statistical methods and results, the 10 second limit was defined a priori. We have taken note of the discursive nature of the statement referred to, moved it to the discussion and amended the text.

Laryngoscopic view is definitely not fair. Anesthesia providers were asked to intubate as soon as possible, but were not asked to provide the best visualization possible. Therefore, providers might not have been focusing on obtaining better visualization, although it would have been possible by maybe even minor improvements.

>> This is a valid comment. Laryngoscopic view was included as an data point to but is not the primary outcome. We have thus included the result for the interest of the reader. We does not necessarily translate to outcome as you correctly point out; we have added a sentence in the discussion reflecting on thi

Discussion:

This needs to be rewritten. Please focus on the current study and discuss the new findings in the context of currently known evidence. Authors do not have to defend this study, or why it is so important to do this study (◊ this should be clear after the introduction).

>> Much of the information provided in the discussion regarding the limitations has been inserted at the request of a prior reviewer, and serves to provide context.

Figure 1. I would suggest to add another picture of the device from the side to indicate the exact curvature of the blade.

>> Side on photo added

Maybe also add the comparison blade, so that reader have an idea, if these blades are comparable.

>> photo added

Figure 2. Although technically correct, this figure does not add anything. I would suggest to delete.

>>Figure 2 (now 3) is a graphical representation of the primary outcome, and is useful for showing both the fact that the devices were equivalent within our *a priori* limits, but also that the CMAC was faster. We feel that it therefore has value.

Table 1. was performed in a sloppy style. There are several empty rows at the bottom.

>> amended

Table 2. Why did the authors choose to report the fremantle score? POGO or even modified C+L would be alternatives

>> Numerous studies in the field have examined the use of different scoring systems for VL view grading. There is consensus that the (modified) Cormack-Lehane grading was neither designed for, nor is fit for this purpose. It is illustrative that all views obtained in this study would be C-L Grade 1. The POGO score is used in some research, but the Freemantle grading has been shown to have a better combination of accuracy and interrater reliability than the POGO (and C-L grade) – we refer the reviewer to the work of O’Loughlin et al in *Anaesthesia* (<https://associationofanaesthetists-publications.onlinelibrary.wiley.com/doi/full/10.1111/anae.13837>)

I am just wondering, but is this device protected by any patent?

>> Yes, the device is patented.

Appendix 3: Abstract Publication TACC

Article

Innovation in low-cost videolaryngoscopy: Intubator V1-Indirect compared with Storz C-MAC in a simulated difficult airway

February 2020 · *Trends in Anaesthesia and Critical Care* 30:e127

DOI:10.1016/j.tacc.2019.12.312

Project: [Context-sensitive airway management in resource-constrained settings](#)

Authors:

Ross Hofmeyr
University of Cape Town

Christella Alphonsus

Christiaan de Villiers
University of Cape Town

Dylan Eave
University of Cape Town

Request full-text

Download citation

Copy link

To read the full-text of this research, you can request a copy directly from the authors.

Citations (1)

Abstract

Videolaryngoscopy (VL) has revolutionised airway management with numerous studies demonstrating its utility to improve airway management and as an emergency device in unanticipated difficult intubation. Availability of VL in all clinical areas has been curtailed by its relative high cost, especially in resource-constrained practice environments such as most of Africa, where research has shown, perioperative outcomes are significantly worse. Advancements in mobile technology, cameras and 3D printing, enabled the design of an innovative low-cost VL, with the vision of making it available to all practitioners. A novel hyper-angulated blade was designed and prototyped using 3D printing. Equipped with a disposable, low-cost, endoscopic camera, it transmits video via a reusable wireless module to a free application on the practitioner's mobile device. The device presents a significant cost reduction to produce, and affordability will enable all practitioners to own a device. In accordance with ADEPT guidelines, a randomised, controlled, crossover equivalence study compared the novel device to an established industry standard (Storz C-MAC with Dēorges blade). A sample size of 97 participants was calculated using PASS software, to power the study to 90% with an alpha error of 2.5% using a clinically determined equivalence limit of 10 seconds mean difference in intubation time. After approval by the UCT Human Research Ethics Committee, 100 skilled intubators, with previous VL experience, received instruction and adequate practice on the new device. They then intubated a Laerdal airway mannikin immobilised to simulate a difficult airway in a randomised order. Video recordings of each device were analysed frame-by-frame to determine time to intubation, measured as the time when the device entered the oral cavity to the time when the ETT black marker moved past the vocal cords. Using the two one-sided test (TOST) method, the mean difference was 4.91 seconds and the 95% confidence limits for the mean difference was between 2.34 e 7.49 seconds, meeting our a priori limits for equivalence, although the C-MAC was faster on average. The innovative low-cost VL was thus equivalent to the industry standard in a simulated difficult airway. Further testing in vivo in a clinical environment is warranted.

Discover the world's research

- 20+ million members
- 135+ million publications
- 700k+ projects [Join for free](#)

Appendix 4: Human Research Ethics Committee approval

UNIVERSITY OF CAPE TOWN
Faculty of Health Sciences
Human Research Ethics Committee

Room E53-46 Old Main Building
Groote Schuur Hospital
Observatory 7925

Telephone [021] 406 6492

Email: sumayah.ariel@uct.ac.za

Website: www.health.uct.ac.za/fhs/research/humanethics/forms

23 November 2018

HREC REF: 770/2018

A/Prof R Hofmeyr

Department of Anaesthesia & Perioperative Medicine
D-23
NGSH

Dear A/Prof Hofmeyr

PROJECT TITLE: LOW COST VIDEO LARYNGOSCOPE, THE INTUBATOR V-1-INDIRECT, COMPARED WITH STORZ-C-MAC WITH D-BLADE IN A SIMULATED DIFFICULT AIRWAY (Masters candidate-Dr C de Villiers)

Thank you for submitting your study to the Faculty of Health Sciences Human Research Ethics Committee (HREC) for review.

It is a pleasure to inform you that the HREC has **formally approved** the above-mentioned study.

Approval is granted for one year until the 30 November 2019.

- The study will require the director of Student affairs approval.
- Please add to the I/C document how confidentiality will be protected.

Please submit a progress form, using the standardised Annual Report Form if the study continues beyond the approval period. Please submit a Standard Closure form if the study is completed within the approval period.

(Forms can be found on our website: www.health.uct.ac.za/fhs/research/humanethics/forms)

We acknowledge that the student: Dr Christiaan de Villiers will also be involved in this study.

Please quote the HREC REF in all your correspondence.

Please note that the ongoing ethical conduct of the study remains the responsibility of the principal investigator.

Please note that for all studies approved by the HREC, the principal investigator **must** obtain appropriate institutional approval, where necessary, before the research may occur.

Yours sincerely

PROFESSOR M BLOCKMAN

Appendix 5: Mannikin study Consent form

CONSENT FORM

TITLE OF THE RESEARCH PROJECT:

Low cost video laryngoscope, the Intubator V1-Indirect, compared with Storz C-MAC with D-Blade in a simulated difficult airway.

PRINCIPAL INVESTIGATOR: Prof R Hofmeyr

CO-INVESTIGATORS: Dr CT de Villiers, Dr C Alphonsus, Dr D Eave

CONTACT DETAILS: Dr CT de Villiers (0825564397, christiaandevilliers@gmail.com)
Dr D Eave (0837765030, doceave@gmail.com)

Dear Practitioner

We are excited to invite you to participate in a research project. Your participation in this research is voluntary and you remain free to withdraw from the research at any point even after you have agreed to take part.

Please take some time to read the information provided below which describes the nature of this research project. You are encouraged to contact the investigators should you have further questions pertaining to this research project.

Your experience, time and cooperation are greatly valued.

This study has been approved by the Health Research Ethics Committee at University of Cape Town, HREC Ref 770/2018 and will be conducted according to the ethical guidelines and principles of the international Declaration of Helsinki, South African Guidelines for Good Clinical Practice and the Medical Research Council (MRC) Ethical Guidelines for Research.

What is being researched?

We are using a high fidelity airway manikin configured with a cervical collar to simulate difficult intubation, to compare the performance of a low cost, hyperangulated, disposable videolaryngoscope device- *Indirect V1 by DisaTech* with that of the *C-MAC D blade by Storz Medical*.

Why have you been invited to participate?

You are a healthcare professional who has experience with direct laryngoscopy in difficult airways. You are also a healthcare professional who may benefit from the availability of another device in your armamentarium for the management of such airways.

How will this research project be conducted?

This research project requires you to practice using both devices, attempt intubation with each device in a predetermined order. Video recorded during the use of the video laryngoscope as well as time taken for intubation attempts with both devices will be captured against your participant number (not your name). Your participation will remain confidential and all data collected will not be traceable back to you. It is estimated that this will take 15 minutes of your time. You will not be remunerated for your involvement in this study.

What is your responsibility if you choose to participate in this research project?

For the duration of the study you will be responsible only for completing all the steps described above to the best of your ability.

Should you have an ethical concern which has not been adequately addressed by the researchers please contact UCT Health Research Ethics Committee at 0214066492. HREC REF 770/2018

Please answer the following questions:

Have you ever used a videolaryngoscope before? Yes _____ No _____

Have you performed more than 200 intubations? Yes _____ No _____

I, _____, the undersigned, hereby consent to take part in the study as described above, and will do so to the best of my ability.

Date: _____ Venue: _____ Signature: _____

PARTICIPANT DATA CAPTURE FORM

Date of Testing: _____

Participant No:	
Which device is tested first?	
Device tested now?	
First or second device:	
Successful:	
Time of Video Capture/ Testing:	

Participant No:	
Which device is tested first?	
Device tested now?	
First or second device:	
Successful:	
Time of Video Capture/ Testing:

TRENDS IN ANAESTHESIA AND CRITICAL CARE

AUTHOR INFORMATION PACK

TABLE OF CONTENTS

• Description	p.1
• Abstracting and Indexing	p.1
• Editorial Board	p.1
• Guide for Authors	p.4

ISSN: 2210-8440

DESCRIPTION

Trends in Anaesthesia and Critical Care provides reviews and comment on highly topical subjects and the latest breakthroughs in basic, clinical and translational research.

Trends in Anaesthesia and Critical Care is an international peer reviewed journal, which provides reviews, opinion articles, lessons from clinical practice, and letters to the editor on highly topical subjects and the latest breakthroughs in basic, clinical and translational research in anaesthesiology and critical care medicine. TACC welcomes also original research articles in airway management.

The journal's objective is to provide a platform for discussion, analysis and debate of topics across a diverse, multidisciplinary audience of basic scientists and clinicians, who share the common goal of understanding clinical practice of **anaesthesiology, critical care**, aspects of related **emergency medicine and resuscitation**, as well as **airway management** with a view to new clinical practice. Communicating emerging concepts and ideas will be facilitated with opinion articles or letters to the editor and help to progress the exciting and evolving field of anaesthesia and critical care. If you are interested in suggesting an article for consideration, please consult the guide for authors of TACC. Authors should note that all articles are subject to peer review.

ABSTRACTING AND INDEXING

Reactions Weekly
Current Abstracts (EBSCO)
Embase
OCLC Contents Alert
Scopus
Emerging Sources Citation Index (ESCI)

EDITORIAL BOARD

Editor-in-Chief

Robert Greif, Bern University Hospital, Bern, Switzerland
Airway management, education, clinical teaching, simulation, faculty development, research ethics

Associate Editors

Javier Belda, University of Valencia, Valencia, Spain

Airway management, Anaesthesiology, Critical care medicine, Mechanical Ventilation, Intensive care medicine

Anjolie Chhabra, Clinical Trials Registry, New Delhi, Delhi, India

Anaesthesiology, Obstetric anaesthesia, Pediatric anaesthesia

Maren Kleine-Brueggene, University Childrens Hospital Zurich, Department of Anaesthesia, Zurich, Switzerland

Paediatric Anaesthesia, Difficult Airway Management, Cardiac Anaesthesia

Massimiliano Sorbello, Policlinico Vittorio Emanuele San Marco University Hospital, Department of Anesthesia and Intensive Care, Catania, Italy

Airway Management, Mechanical Ventilation, tracheostomy, Critical Care, Pulmonary function

Editorial Board

Edwin Seet Chuen Ping, Khoo Teck Puat Hospital, Singapore, Singapore

Airway management, VL, SGD, Obstructive sleep apnea

Edoardo De Robertis, University of Perugia, Department of Surgical and Biomedical Sciences, Perugia, Italy

Ventilation, monitoring, infection, critical emergency medicine, obstetric anaesthesia, robotic surgery, bleeding

Stefano Falcetta, University Hospital of Ancona Umberto I G M Lancisi G Salesi, Ancona, Italy

Airway Management, Ultrasound

Matthias Haengg, University of Bern, University Hospital Bern, Department of Intensive Care Medicine, Bern, Switzerland

Critical Care Medicine, Resuscitation, Cardiac Arrest

Ross Hofmeyr, University of Cape Town Department of Anaesthesia and Perioperative Medicine, Cape Town, South Africa

Airway management, thoracic anaesthesia, point-of-care testing, prehospital anaesthesia and critical care retrieval, wilderness medicine, extreme physiology

Nian Chih Hwang, Singapore General Hospital's Department of Anaesthesiology, Singapore, Singapore

Cardiac anaesthesia, medical education

Bruce MacIver, Stanford University School of Medicine Department of Anesthesiology Perioperative and Pain Medicine, Stanford, California, United States of America

Anaesthetic mechanisms of action, central nervous system

Sheila Nainan Myatra, Tata Memorial Hospital Department of Anaesthesia, Mumbai, India

Airway Management, Hemodynamic Monitoring, ICU sedation, Prevention of Nosocomial Infections and End of Life Care

Sabine Nabecker, University of Toronto Department of Anesthesia, Toronto, Ontario, Canada

Airway management, medical education and simulation, qualitative studies

Maria Isabel Real, 12th of October University Hospital, Department of Critical Care, Pain Therapy and Anesthesiology, Madrid, Spain

Organ Transplantation, Pulmonary Hypertension, Echocardiography, Thoracic Surgery and Airway Management

Stefano Romagnoli, University of Florence, Florence University Hospital, Department of Anesthesia and Critical Care, Firenze, Italy

Sepsis, Hemodynamic Monitoring, Acute Kidney Injury, Renal Replacement Therapy, Sedation, Depth of Anesthesia and Sedation pEEG-based monitoring, Sleep in critically ill patients

Patrick Schoettker, Lausanne University Hospital Department of Anaesthesiology, Lausanne, Switzerland

Airway, technology, medical device

Amit Shah, Shubhechha Multispeciality Hospital, Vadodra, India

Airway management, Lowflow anaesthesia and machines, Laparoscopic anaesthesia, Neuroanaesthesia, Onco anaesthesia

Marina Soro, University Clinic Hospital, Department of Anesthesia and Critical Care, Valencia, Spain

Perioperative Care, Respiratory system, Mechanical ventilation

Zoltán Pál Szűcs, Péterfy Street Municipal Hospital and Trauma Center, Anesthesiology and Intensive Care, Budapest, Hungary

Airway management, trauma care, brain injuries

Thomas Uray, Medical University of Vienna Department of Emergency Medicine, Wien, Austria

Cardiac arrest, therapeutic hypothermia, prehospital emergency medicine

Felipe Urdaneta, University of Florida Department of Anesthesiology, Gainesville, Florida, United States of America

Anesthesia, Airway management, Patient safety

David Weidenauer, Medical University of Vienna, Vienna, Austria

Resuscitation, Medical Education, Cardiology

EAMS Section Editors

Paula Chiesa, University of Buenos Aires, Buenos Aires, Argentina

Airway Management, Medical Education, Pharmacology

Ana Isabel Pereira, Hospital Centre of Vila Nova de Gaia Espinho, Department of Anaesthesiology, Vila Nova de Gaia, Portugal

Anaesthesiology, Airway Management, Education, Emergency Medicine

Kemal Tolga Saracoglu, Istanbul Doctor Lufti Kirdar Kartal Training and Research Hospital, Department of Anesthesiology and Intensive Care, Istanbul, Turkey

Airway management

Editorial Assistant

John Prince

GUIDE FOR AUTHORS

Submission checklist

You can use this list to carry out a final check of your submission before you send it to the journal for review. Please check the relevant section in this Guide for Authors for more details.

Ensure that the following items are present:

One author has been designated as the corresponding author with contact details:

- E-mail address
- Full postal address

All necessary files have been uploaded:

Manuscript:

- Include keywords
- All figures (include relevant captions)
- All tables (including titles, description, footnotes)
- Ensure all figure and table citations in the text match the files provided
- Indicate clearly if color should be used for any figures in print

Graphical Abstracts / Highlights files (where applicable)

Supplemental files (where applicable)

Further considerations

- Manuscript has been 'spell checked' and 'grammar checked'
- All references mentioned in the Reference List are cited in the text, and vice versa
- Permission has been obtained for use of copyrighted material from other sources (including the Internet)
- A competing interests statement is provided, even if the authors have no competing interests to declare
- Journal policies detailed in this guide have been reviewed
- Referee suggestions and contact details provided, based on journal requirements

For further information, visit our [Support Center](#).

Article Types

Lessons from Clinical Practice. These are an opportunity to publish an interesting case or case series which you have encountered. They should not be regarded as the traditional 'case report' because they are more detailed than that. Authors are asked to submit an interesting case or clinical lesson for consideration. Each will be considered individually and only those with a real interesting clinical lesson will be regarded as possibly suitable for publication. If it is accepted then it will be sent on to an expert who will write an accompanying piece providing a critique or further information concerning the situation. It should not be used as an opportunity to try to get random case reports published, to promote original work, or to introduce new data.

Lessons should be interesting and novel. They should aim to tell the reader about an interesting case or situation from which real lessons can be learnt. Very rare cases would only be suitable for consideration if they illustrated an important lesson. The format should be to start with a brief introduction (contextual background) followed by an account of what happened or was observed. A discussion section should then be included. The lesson should be referenced accordingly.

Remember that you are writing for a wide audience - not only for readers in your own field. You can assume that your audience is competent in the basic language of the subject, but may require explanation or definition of technical terms, concepts and assumptions specific to your topic.

Length: 500-800 words, 12 references, 1 figure/box/table.

Authors names: No more than 3 authors. Full contact details for all authors (indicate corresponding author).

Abstract: Briefly explain the necessary background and encapsulate the take-home message for a non-specialist readership. 100 to 120 words.

Figures (optional): Should always have a short, explanatory title (as well as legend). Legends must fully explain the figure without reference to the text.

Tables (optional): Require a single-sentence title but no legend. Abbreviations (if not defined in the main text) and full explanations should be footnoted (using letters).

Text Boxes (optional): Ideal for providing explanations of basic concepts or theories, giving detailed mechanisms or discussing case studies. Text Boxes can occasionally contain small figures and tables. 200 words max per Box (references to be listed in main reference list only).

The commentary will be written by an expert in the field and will have the same limits as those given above, i.e. maximum of 500-800 words, one figure or table and 12 references.

Letters to the Editor are reserved for genuine synthesis only.

Letters should have a journalistic flavour: aim to tell a story and communicate your enthusiasm for the subject. Start with some very brief contextual background, and be sure to alert the reader to the main message of your letter before the end of the first paragraph.

You are writing for a wide audience, not only for readers in your own field. You can assume that your audience is competent in the basic language of the subject, but may require explanation or definition of technical terms, concepts and assumptions specific to your topic. Avoid jargon, but do not oversimplify, be accurate and precise throughout. Use sections and subheadings to lead your reader through the discussion.

Length: 750-800 words, 12 references, 1 figure/box/table.

Authors names: No more than 3 authors. Include full contact details for all authors (indicate corresponding author).

Figures (optional): Should always have a short, explanatory title (as well as legend). Legends must fully explain the figure without reference to the text.

Tables (optional): Require a single-sentence title but no legend. Abbreviations (if not defined in the main text) and full explanations should be footnoted (using letters).

Text Boxes (optional): Ideal for providing explanations of basic concepts or theories, giving detailed mechanisms or discussing case studies. Text Boxes can occasionally contain small figures and tables. 200 words max per Box. References to be listed in main reference list only.

TACC Opinions should present a personal viewpoint on a research-related topic, rather than a review of this topic. The aim should be to stimulate debate or new research, cover controversial topics, or provide a new framework for, or interpretation of, an old problem or current issue, or speculate on the implications of recent research and clinical advances in health care infrastructure concerns at national or international level.

Articles that outline recent advances in a field rather than give a strong opinion on them are not suitable for this section of the journal. Our audience ranges from student to professor, and so articles must be accessible to a wide readership.

TACC Opinions must not include unpublished original data. Although subjective, an opinion article should not be used to dwell excessively on the author's own research or excessively criticise others', except where criticism is constructive.

Title: Should be short and enticing (no more than eight words).

Authors: No more than 3 authors. Full contact details for all authors (indicate corresponding author).

Abstract: Briefly explain the necessary background and encapsulate the take-home message for a non-specialist readership. Please emphasize the opinion you are putting forward. 100 to 120 words. Do not include reference citations.

Opinion structure: Please start with an accessible introduction outlining the timeliness, importance and rationale behind your article (why the subject is important, why now). Restate your opinion clearly. Use concise logical subheadings and provide clear links between sections. Finish with clearly stated conclusions, including an indication of future directions.

Length: 2500-3500 words: This limit does not include text in boxes, abstract, references, tables or figure legends. **Figures:** Should always have a short, explanatory title to preface the legend. Legends must fully explain the figure without reference to the text.

Tables: Require a single-sentence title but no legend. Abbreviations (if not defined in the main text) and full explanations should be footnoted (using letters).

Tables: Require a single-sentence title but no legend. Abbreviations (if not defined in the main text) and full explanations should be footnoted (using letters).

Text Boxes: Ideal for providing explanations of basic concepts or theories, giving detailed mechanisms or discussing case studies. Text Boxes can occasionally contain small figures and tables. 400 words max per Box (refs. to be listed in main reference list only: see checklist). No more than 3 boxes per article. Outstanding questions: Can be summarized in a box (not included in box count).

Glossary Box: Used for specialist language (not included in box count).

References: The reference list should not be exhaustive - simply alert the readers to the key literature on which your opinion is based. Reviews should be cited if necessary to refer to older data. Up to 50 references.

TACC Airway Management Original Articles welcome all areas of airway management from anesthesia, critical care, and emergency medicine in- and outside of hospitals, to teaching and training issues in airway management, as well as emerging concepts and ideas about equipment for airway management in human or manikin studies, as long as the authors can demonstrate their clinical or educational relevance of the issues investigated and described in their articles with the final aim to improved safe patient care. Our audience ranges from student to professor, and so articles must be accessible to a wider readership.

Authors should note that all articles are subject to a rigorous peer review process and commonly accepted ethical rules, like written informed consent by the patients or study participants as well as ethic committee approval is mandatory. Reviewers will have to answer for each article the following questions before they can finally recommend a suggestion to the editor about acceptance of an article to be published: Was the subject addressed in the submitted article worthy of investigation and publishing? Is the information presented new? Are the conclusions drawn by the authors supported by their data? Is the manuscript appropriate for the EAMS section? Is the organization of the manuscript appropriate? Are the figures, tables and supplementary data appropriate? And finally was the conduct of the study ethically acceptable?

To help you prepare your article, please take careful note of the following guidelines (especially limits and 'Article structure'). A more detailed formatting checklist follows these instructions.

Title: Titles should be short and enticing.

Authors names: No more than 7 authors - Full contact details for all authors (indicate corresponding author).

Structured Abstract with headings to include: Background, Aim, Methods, Results, Conclusion: Briefly explain the necessary background and encapsulate the take-home message. Please emphasize the recent developments that make your research timely. 250 to 500 words.

Article structure: Introduction: please indicate the timeliness and rationale for your article (why the subject is important; why now). Use concise logical subheadings and provide clear links between sections. Please end with a brief summary of your article, a strong take-home message and include a clear indication of future research.

Length: 3500 words: this limit does not include text in boxes, tables, figure legends abstract or references

Figures: Should always have a short, explanatory title to preface the legend. Legends must fully explain the figure without reference to the text.

Tables: Require a single-sentence title but no legend. Abbreviations (if not defined in the main text) and full explanations should be footnoted (using letters).

Text Boxes: Ideal for providing explanations of basic concepts or theories, giving detailed mechanisms or discussing case studies. Text Boxes can occasionally contain small figures and tables. 400 words max per Box (refs. to be listed in main reference list only: see checklist). No more than 4 boxes per article.

Outstanding questions: Can be summarized in a box (not included in box count).

Glossary Box: Used for specialist language (not included in box count).

References: 30-35 references should not be exceeded.

BEFORE YOU BEGIN

Ethics in publishing

Please see our information pages on [Ethics in publishing](#) and [Ethical guidelines for journal publication](#).

Studies in humans and animals

If the work involves the use of human subjects, the author should ensure that the work described has been carried out in accordance with [The Code of Ethics of the World Medical Association](#) (Declaration of Helsinki) for experiments involving humans. The manuscript should be in line with the

Recommendations for the Conduct, Reporting, Editing and Publication of Scholarly Work in Medical Journals and aim for the inclusion of representative human populations (sex, age and ethnicity) as per those recommendations. The terms **sex** and **gender** should be used correctly.

Authors should include a statement in the manuscript that informed consent was obtained for experimentation with human subjects. The privacy rights of human subjects must always be observed.

All animal experiments should comply with the **ARRIVE guidelines** and should be carried out in accordance with the U.K. Animals (Scientific Procedures) Act, 1986 and associated guidelines, **EU Directive 2010/63/EU for animal experiments**, or the National Institutes of Health guide for the care and use of Laboratory animals (NIH Publications No. 8023, revised 1978) and the authors should clearly indicate in the manuscript that such guidelines have been followed. The sex of animals must be indicated, and where appropriate, the influence (or association) of sex on the results of the study.

Informed consent and patient details

Studies on patients or volunteers require ethics committee approval and informed consent, which should be documented in the paper. Appropriate consents, permissions and releases must be obtained where an author wishes to include case details or other personal information or images of patients and any other individuals in an Elsevier publication. Written consents must be retained by the author but copies should not be provided to the journal. Only if specifically requested by the journal in exceptional circumstances (for example if a legal issue arises) the author must provide copies of the consents or evidence that such consents have been obtained. For more information, please review the **Elsevier Policy on the Use of Images or Personal Information of Patients or other Individuals**. Unless you have written permission from the patient (or, where applicable, the next of kin), the personal details of any patient included in any part of the article and in any supplementary materials (including all illustrations and videos) must be removed before submission.

Reporting Guidelines

The Editor requires that manuscripts adhere to recognized reporting guidelines relevant to the research design used and require authors to submit a checklist verifying that essential elements have been reported for all primary research and systematic reviews. Reporting guidelines endorsed by the journal are listed below:

Observational cohort, case control and cross sectional studies - STROBE - Strengthening the Reporting of Observational Studies in Epidemiology, <http://www.equator-network.org/reporting-guidelines/strobe/> *Systematic Reviews* - PROSPERO - International register of systematic reviews, <https://www.crd.york.ac.uk/prospere/#aboutregpage> *Qualitative studies* - COREQ - Consolidated criteria for reporting qualitative research, <http://www.equator-network.org/reporting-guidelines/coreq> *Quasi-experimental/non-randomised evaluations* - TREND - Transparent Reporting of Evaluations with Non-randomized Designs, <http://www.cdc.gov/trendstatement/> *Case Reports* - CARE Guidelines-<https://www.care-statement.org/> *Randomised (and quasi-randomised) controlled trial* - CONSORT - Consolidated Standards of Reporting Trials, <http://www.equator-network.org/reporting-guidelines/consort/> *Animal Research* - ARRIVE- <https://www.nc3rs.org.uk/arrive-guidelines> *Study of Diagnostic accuracy/assessment scale* - STARD - Standards for the Reporting of Diagnostic Accuracy Studies, <http://www.equator-network.org/reporting-guidelines/stard/> *Systematic Review of Controlled Trials* - PRISMA - Preferred Reporting Items for Systematic Reviews and Meta-Analyses, <http://www.equator-network.org/reporting-guidelines/prisma/> *SRQR checklist is for qualitative studies* <http://www.equator-network.org/reporting-guidelines/srqr/> *AGREE checklist is for clinical guidelines* type of articles <http://www.equator-network.org/reporting-guidelines/the-agree-reporting-checklist-a-tool-to-improve-reporting-of-clinical-practice-guidelines/>

You are required to adhere to these guidelines (or a suitable recognized alternative) and to submit a completed checklist from the reporting guideline to assist the editors and reviewers of your paper. You can search for the correct guideline for your study using the tools provided by the EQUATOR network: <http://www.equator-network.org/> The guideline used must be indicated in the **Author Checklist** and the completed Standards of Reporting Checklist must also be included in your submission.

Declaration of interest

All authors must disclose any financial and personal relationships with other people or organizations that could inappropriately influence (bias) their work. Examples of potential competing interests include employment, consultancies, stock ownership, honoraria, paid expert testimony, patent applications/registrations, and grants or other funding. **Authors must disclose any interests in two places:**

1. A summary Conflict of interest Statement in a separate Word document. This should include any detail that may be considered a conflict of interest for all authors. If there are no interests to declare for any of the authors then please state this: 'Declarations of interest: none'. **This summary statement will be ultimately published if the article is accepted so needs to be in an editable Word format. Please ensure it includes enough information for researchers as they will not have access to the forms.**

2. Detailed disclosures within separate Conflict of Interest / ICMJE forms. **It is important for potential interests to be declared in both places and that the information matches.** Please [click here for more information](#). These form(s) will not be published and are for the review process and the journal's official records only.

Submission declaration and verification

Submission of an article implies that the work described has not been published previously (except in the form of an abstract, a published lecture or academic thesis, see '[Multiple, redundant or concurrent publication](#)' for more information), that it is not under consideration for publication elsewhere, that its publication is approved by all authors and tacitly or explicitly by the responsible authorities where the work was carried out, and that, if accepted, it will not be published elsewhere in the same form, in English or in any other language, including electronically without the written consent of the copyright-holder. To verify originality, your article may be checked by the originality detection service [Crossref Similarity Check](#).

Preprints

Please note that [preprints](#) can be shared anywhere at any time, in line with Elsevier's [sharing policy](#). Sharing your preprints e.g. on a preprint server will not count as prior publication (see '[Multiple, redundant or concurrent publication](#)' for more information).

Use of inclusive language

Inclusive language acknowledges diversity, conveys respect to all people, is sensitive to differences, and promotes equal opportunities. Content should make no assumptions about the beliefs or commitments of any reader; contain nothing which might imply that one individual is superior to another on the grounds of age, gender, race, ethnicity, culture, sexual orientation, disability or health condition; and use inclusive language throughout. Authors should ensure that writing is free from bias, stereotypes, slang, reference to dominant culture and/or cultural assumptions. We advise to seek gender neutrality by using plural nouns ("clinicians, patients/clients") as default/wherever possible to avoid using "he, she," or "he/she." We recommend avoiding the use of descriptors that refer to personal attributes such as age, gender, race, ethnicity, culture, sexual orientation, disability or health condition unless they are relevant and valid. These guidelines are meant as a point of reference to help identify appropriate language but are by no means exhaustive or definitive.

Author contributions

For transparency, we encourage authors to submit an author statement file outlining their individual contributions to the paper using the relevant CRediT roles: Conceptualization; Data curation; Formal analysis; Funding acquisition; Investigation; Methodology; Project administration; Resources; Software; Supervision; Validation; Visualization; Roles/Writing - original draft; Writing - review & editing. Authorship statements should be formatted with the names of authors first and CRediT role(s) following. [More details and an example](#)

Authors are expected to consider carefully the list and order of authors before submitting their manuscript and provide the definitive list of authors at the time of the original submission. Any addition, deletion or rearrangement of author names in the authorship list should be made only before the manuscript has been accepted and only if approved by the journal Editor.

To request such a change, the Editor must receive the following from the corresponding author:

- (a) the reason for the change in author list and
- (b) written confirmation (e-mail, letter) from all authors that they agree with the addition, removal or rearrangement.

In the case of addition or removal of authors, this includes confirmation from the author being added or removed. Only in exceptional circumstances will the Editor consider the addition, deletion or rearrangement of authors after the manuscript has been accepted.

While the Editor considers the request, publication of the manuscript will be suspended. If the manuscript has already been published in an online issue, any requests approved by the Editor will result in a corrigendum.

Reporting clinical trials

Randomized controlled trials should be presented according to the CONSORT guidelines. At manuscript submission, authors must provide the CONSORT checklist accompanied by a flow diagram that illustrates the progress of patients through the trial, including recruitment, enrollment, randomization, withdrawal and completion, and a detailed description of the randomization procedure. The [CONSORT checklist and template flow diagram](#) are available online.

Registration of clinical trials

Registration in a public trials registry is a condition for publication of clinical trials in this journal in accordance with [International Committee of Medical Journal Editors](#) recommendations. Trials must register at or before the onset of patient enrolment. The clinical trial registration number should be included at the end of the abstract of the article. A clinical trial is defined as any research study that prospectively assigns human participants or groups of humans to one or more health-related interventions to evaluate the effects of health outcomes. Health-related interventions include any intervention used to modify a biomedical or health-related outcome (for example drugs, surgical procedures, devices, behavioural treatments, dietary interventions, and process-of-care changes). Health outcomes include any biomedical or health-related measures obtained in patients or participants, including pharmacokinetic measures and adverse events. Purely observational studies (those in which the assignment of the medical intervention is not at the discretion of the investigator) will not require registration.

Article transfer service

This journal is part of our Article Transfer Service. This means that if the Editor feels your article is more suitable in one of our other participating journals, then you may be asked to consider transferring the article to one of those. If you agree, your article will be transferred automatically on your behalf with no need to reformat. Please note that your article will be reviewed again by the new journal. [More information.](#)

Copyright

Upon acceptance of an article, authors will be asked to complete a 'Journal Publishing Agreement' (see [more information](#) on this). An e-mail will be sent to the corresponding author confirming receipt of the manuscript together with a 'Journal Publishing Agreement' form or a link to the online version of this agreement.

Subscribers may reproduce tables of contents or prepare lists of articles including abstracts for internal circulation within their institutions. [Permission](#) of the Publisher is required for resale or distribution outside the institution and for all other derivative works, including compilations and translations. If excerpts from other copyrighted works are included, the author(s) must obtain written permission from the copyright owners and credit the source(s) in the article. Elsevier has [preprinted forms](#) for use by authors in these cases.

For gold open access articles: Upon acceptance of an article, authors will be asked to complete a 'License Agreement' ([more information](#)). Permitted third party reuse of gold open access articles is determined by the author's choice of [user license](#).

Author rights

As an author you (or your employer or institution) have certain rights to reuse your work. [More information.](#)

Elsevier supports responsible sharing

Find out how you can [share your research](#) published in Elsevier journals.

Role of the funding source

You are requested to identify who provided financial support for the conduct of the research and/or preparation of the article and to briefly describe the role of the sponsor(s), if any, in study design; in the collection, analysis and interpretation of data; in the writing of the report; and in the decision to submit the article for publication. If the funding source(s) had no such involvement then this should be stated.

Open access

Please visit our [Open Access](#) page for more information.

Elsevier Researcher Academy

[Researcher Academy](#) is a free e-learning platform designed to support early and mid-career researchers throughout their research journey. The "Learn" environment at Researcher Academy offers several interactive modules, webinars, downloadable guides and resources to guide you through the process of writing for research and going through peer review. Feel free to use these free resources to improve your submission and navigate the publication process with ease.

Language (usage and editing services)

Please write your text in good English (American or British usage is accepted, but not a mixture of these). Authors who feel their English language manuscript may require editing to eliminate possible grammatical or spelling errors and to conform to correct scientific English may wish to use the [English Language Editing service](#) available from Elsevier's Author Services.

Submission

Our online submission system guides you stepwise through the process of entering your article details and uploading your files. The system converts your article files to a single PDF file used in the peer-review process. Editable files (e.g., Word, LaTeX) are required to typeset your article for final publication. All correspondence, including notification of the Editor's decision and requests for revision, is sent by e-mail.

PREPARATION

Peer review

This journal operates a single anonymized review process. All contributions are typically sent to a minimum of two independent expert reviewers to assess the scientific quality of the paper. The Editor is responsible for the final decision regarding acceptance or rejection of articles. The Editor's decision is final. Editors are not involved in decisions about papers which they have written themselves or have been written by family members or colleagues or which relate to products or services in which the editor has an interest. Any such submission is subject to all of the journal's usual procedures, with peer review handled independently of the relevant editor and their research groups. [More information on types of peer review.](#)

Submission by an Editor

A paper submitted by an editor or board member (including submission as a co-author) will be handled by one of the other Senior Editors who will manage the peer review process. The Journal's submission system does not allow Editors to manage their own papers in the system.

Article structure

Subdivision - numbered sections

Divide your article into clearly defined and numbered sections. Subsections should be numbered 1.1 (then 1.1.1, 1.1.2, ...), 1.2, etc. (the abstract is not included in section numbering). Use this numbering also for internal cross-referencing: do not just refer to 'the text'. Any subsection may be given a brief heading. Each heading should appear on its own separate line.

Introduction

State the objectives of the work and provide an adequate background, avoiding a detailed literature survey or a summary of the results.

Essential title page information

- **Title.** Concise and informative. Titles are often used in information-retrieval systems. Avoid abbreviations and formulae where possible.
- **Author names and affiliations.** Please clearly indicate the given name(s) and family name(s) of each author and check that all names are accurately spelled. You can add your name between parentheses in your own script behind the English transliteration. Present the authors' affiliation addresses (where the actual work was done) below the names. Indicate all affiliations with a lower-

case superscript letter immediately after the author's name and in front of the appropriate address. Provide the full postal address of each affiliation, including the country name and, if available, the e-mail address of each author.

• **Corresponding author.** Clearly indicate who will handle correspondence at all stages of refereeing and publication, also post-publication. This responsibility includes answering any future queries about Methodology and Materials. **Ensure that the e-mail address is given and that contact details are kept up to date by the corresponding author.**

• **Present/permanent address.** If an author has moved since the work described in the article was done, or was visiting at the time, a 'Present address' (or 'Permanent address') may be indicated as a footnote to that author's name. The address at which the author actually did the work must be retained as the main, affiliation address. Superscript Arabic numerals are used for such footnotes.

Highlights

Highlights are optional yet highly encouraged for this journal, as they increase the discoverability of your article via search engines. They consist of a short collection of bullet points that capture the novel results of your research as well as new methods that were used during the study (if any). Please have a look at the examples here: [example Highlights](#).

Highlights should be submitted in a separate editable file in the online submission system. Please use 'Highlights' in the file name and include 3 to 5 bullet points (maximum 85 characters, including spaces, per bullet point).

Abstract

A concise and factual abstract is required. The abstract should state briefly the purpose of the research, the principal results and major conclusions. An abstract is often presented separately from the article, so it must be able to stand alone. For this reason, References should be avoided, but if essential, then cite the author(s) and year(s). Also, non-standard or uncommon abbreviations should be avoided, but if essential they must be defined at their first mention in the abstract itself.

Graphical abstract

Although a graphical abstract is optional, its use is encouraged as it draws more attention to the online article. The graphical abstract should summarize the contents of the article in a concise, pictorial form designed to capture the attention of a wide readership. Graphical abstracts should be submitted as a separate file in the online submission system. Image size: Please provide an image with a minimum of 531 × 1328 pixels (h × w) or proportionally more. The image should be readable at a size of 5 × 13 cm using a regular screen resolution of 96 dpi. Preferred file types: TIFF, EPS, PDF or MS Office files. You can view [Example Graphical Abstracts](#) on our information site.

Authors can make use of Elsevier's [Illustration Services](#) to ensure the best presentation of their images and in accordance with all technical requirements.

Keywords

Immediately after the abstract, provide a maximum of 6 keywords, using American spelling and avoiding general and plural terms and multiple concepts (avoid, for example, 'and', 'of'). Be sparing with abbreviations: only abbreviations firmly established in the field may be eligible. These keywords will be used for indexing purposes.

Abbreviations

Define abbreviations that are not standard in this field in a footnote to be placed on the first page of the article. Such abbreviations that are unavoidable in the abstract must be defined at their first mention there, as well as in the footnote. Ensure consistency of abbreviations throughout the article.

Acknowledgements

Collate acknowledgements in a separate section at the end of the article before the references and do not, therefore, include them on the title page, as a footnote to the title or otherwise. List here those individuals who provided help during the research (e.g., providing language help, writing assistance or proof reading the article, etc.).

Formatting of funding sources

List funding sources in this standard way to facilitate compliance to funder's requirements:

Funding: This work was supported by the National Institutes of Health [grant numbers xxxx, yyyy]; the Bill & Melinda Gates Foundation, Seattle, WA [grant number zzzz]; and the United States Institutes of Peace [grant number aaaa].

It is not necessary to include detailed descriptions on the program or type of grants and awards. When funding is from a block grant or other resources available to a university, college, or other research institution, submit the name of the institute or organization that provided the funding.

If no funding has been provided for the research, please include the following sentence:

This research did not receive any specific grant from funding agencies in the public, commercial, or not-for-profit sectors.

Footnotes

Footnotes should be used sparingly. Number them consecutively throughout the article. Many word processors can build footnotes into the text, and this feature may be used. Otherwise, please indicate the position of footnotes in the text and list the footnotes themselves separately at the end of the article. Do not include footnotes in the Reference list.

Artwork

Electronic artwork

General points

- Make sure you use uniform lettering and sizing of your original artwork.
- Embed the used fonts if the application provides that option.
- Aim to use the following fonts in your illustrations: Arial, Courier, Times New Roman, Symbol, or use fonts that look similar.
- Number the illustrations according to their sequence in the text.
- Use a logical naming convention for your artwork files.
- Provide captions to illustrations separately.
- Size the illustrations close to the desired dimensions of the published version.
- Submit each illustration as a separate file.
- Ensure that color images are accessible to all, including those with impaired color vision.

A detailed [guide on electronic artwork](#) is available.

You are urged to visit this site; some excerpts from the detailed information are given here.

Formats

If your electronic artwork is created in a Microsoft Office application (Word, PowerPoint, Excel) then please supply 'as is' in the native document format.

Regardless of the application used other than Microsoft Office, when your electronic artwork is finalized, please 'Save as' or convert the images to one of the following formats (note the resolution requirements for line drawings, halftones, and line/halftone combinations given below):

EPS (or PDF): Vector drawings, embed all used fonts.

TIFF (or JPEG): Color or grayscale photographs (halftones), keep to a minimum of 300 dpi.

TIFF (or JPEG): Bitmapped (pure black & white pixels) line drawings, keep to a minimum of 1000 dpi.

TIFF (or JPEG): Combinations bitmapped line/half-tone (color or grayscale), keep to a minimum of 500 dpi.

Please do not:

- Supply files that are optimized for screen use (e.g., GIF, BMP, PICT, WPG); these typically have a low number of pixels and limited set of colors;
- Supply files that are too low in resolution;
- Submit graphics that are disproportionately large for the content.

Colour Artwork

Please make sure that artwork files are in an acceptable format (TIFF (or JPEG), EPS (or PDF), or MS Office files) and with the correct resolution. If you were commissioned to write your article, and together with your accepted article you submit usable color figures, then Elsevier will ensure that at no additional charge these figures will appear in color on the Web (e.g., ScienceDirect and other sites) in addition to color reproduction in print. If you were not commissioned then **for color reproduction in print, you will receive information regarding the costs from Elsevier after receipt of your accepted article.** Please indicate your preference for color: in print or on the Web only. For further information on the preparation of electronic artwork, please see <https://www.elsevier.com/artworkinstructions>.

Please note: Because of technical complications which can arise by converting color figures to 'gray scale' (for the printed version should you not opt for color in print) please submit in addition usable black and white versions of all the color illustrations.

Illustration services

Elsevier's Author Services offers Illustration Services to authors preparing to submit a manuscript but concerned about the quality of the images accompanying their article. Elsevier's expert illustrators can produce scientific, technical and medical-style images, as well as a full range of charts, tables and graphs. Image 'polishing' is also available, where our illustrators take your image(s) and improve them to a professional standard. Please visit the website to find out more.

Figure captions

Ensure that each illustration has a caption. Supply captions separately, not attached to the figure. A caption should comprise a brief title (**not** on the figure itself) and a description of the illustration. Keep text in the illustrations themselves to a minimum but explain all symbols and abbreviations used.

Tables

Please submit tables as editable text and not as images. Tables can be placed either next to the relevant text in the article, or on separate page(s) at the end. Number tables consecutively in accordance with their appearance in the text and place any table notes below the table body. Be sparing in the use of tables and ensure that the data presented in them do not duplicate results described elsewhere in the article. Please avoid using vertical rules and shading in table cells.

References

Citation in text

Please ensure that every reference cited in the text is also present in the reference list (and vice versa). Any references cited in the abstract must be given in full. Unpublished results and personal communications are not recommended in the reference list, but may be mentioned in the text. If these references are included in the reference list they should follow the standard reference style of the journal (including the article title, authors' names and the journal title where applicable) and should include a substitution of the publication date with either 'Unpublished results' or 'Personal communication'. Citation of a reference as 'in press' implies that the item has been accepted for publication and this can be placed at the end of the reference (e.g. 2018, in press).

Web references

As a minimum, the full URL should be given and the date when the reference was last accessed. Any further information, if known (DOI, author names, dates, reference to a source publication, etc.), should also be given. Web references can be listed separately (e.g., after the reference list) under a different heading if desired, or can be included in the reference list.

Data references

This journal encourages you to cite underlying or relevant datasets in your manuscript by citing them in your text and including a data reference in your Reference List. Data references should include the following elements: author name(s), dataset title, data repository, version (where available), year, and global persistent identifier. Add [dataset] immediately before the reference so we can properly identify it as a data reference. The [dataset] identifier will not appear in your published article.

References in a special issue

Please ensure that the words 'this issue' are added to any references in the list (and any citations in the text) to other articles in the same Special Issue.

Reference style

Text: Indicate references by number(s) in square brackets in line with the text. The actual authors can be referred to, but the reference number(s) must always be given.

Example: '.... as demonstrated [3,6]. Barnaby and Jones [8] obtained a different result'

List: Number the references (numbers in square brackets) in the list in the order in which they appear in the text.

Examples:

Reference to a journal publication:

[1] J. van der Geer, J.A.J. Hanraads, R.A. Lupton, The art of writing a scientific article, J. Sci. Commun. 163 (2010) 51–59. <https://doi.org/10.1016/j.Sc.2010.00372>.

Reference to a journal publication with an article number:

[2] J. van der Geer, J.A.J. Hanraads, R.A. Lupton, 2018. The art of writing a scientific article. Heliyon. 19, e00205. <https://doi.org/10.1016/j.heliyon.2018.e00205>.

Reference to a book:

[3] W. Strunk Jr., E.B. White, The Elements of Style, fourth ed., Longman, New York, 2000.

Reference to a chapter in an edited book:

[4] G.R. Mettam, L.B. Adams, How to prepare an electronic version of your article, in: B.S. Jones, R.Z. Smith (Eds.), *Introduction to the Electronic Age*, E-Publishing Inc., New York, 2009, pp. 281–304.

Reference to a website:

[5] Cancer Research UK, Cancer statistics reports for the UK. <http://www.cancerresearchuk.org/aboutcancer/statistics/cancerstatsreport/>, 2003 (accessed 13 March 2003).

Reference to a dataset:

[dataset] [6] M. Oguro, S. Imahiro, S. Saito, T. Nakashizuka, Mortality data for Japanese oak wilt disease and surrounding forest compositions, Mendeley Data, v1, 2015. <https://doi.org/10.17632/xwj98nb39r.1>.

Journal abbreviations source

Journal names should be abbreviated according to the [List of Title Word Abbreviations](#).

Revised Submissions

Use of word processing software

Regardless of the file format of the original submission, at revision you must provide us with an editable file of the entire article. Keep the layout of the text as simple as possible. Most formatting codes will be removed and replaced on processing the article. The electronic text should be prepared in a way very similar to that of conventional manuscripts (see also the [Guide to Publishing with Elsevier](#)). See also the section on Electronic artwork.

To avoid unnecessary errors you are strongly advised to use the 'spell-check' and 'grammar-check' functions of your word processor.

Video

Elsevier accepts video material and animation sequences to support and enhance your scientific research. Authors who have video or animation files that they wish to submit with their article are strongly encouraged to include links to these within the body of the article. This can be done in the same way as a figure or table by referring to the video or animation content and noting in the body text where it should be placed. All submitted files should be properly labeled so that they directly relate to the video file's content. In order to ensure that your video or animation material is directly usable, please provide the file in one of our recommended file formats with a preferred maximum size of 150 MB per file, 1 GB in total. Video and animation files supplied will be published online in the electronic version of your article in Elsevier Web products, including [ScienceDirect](#). Please supply 'stills' with your files: you can choose any frame from the video or animation or make a separate image. These will be used instead of standard icons and will personalize the link to your video data. For more detailed instructions please visit our [video instruction pages](#). Note: since video and animation cannot be embedded in the print version of the journal, please provide text for both the electronic and the print version for the portions of the article that refer to this content.

Data visualization

Include interactive data visualizations in your publication and let your readers interact and engage more closely with your research. Follow the instructions [here](#) to find out about available data visualization options and how to include them with your article.

Supplementary material

Supplementary material such as applications, images and sound clips, can be published with your article to enhance it. Submitted supplementary items are published exactly as they are received (Excel or PowerPoint files will appear as such online). Please submit your material together with the article and supply a concise, descriptive caption for each supplementary file. If you wish to make changes to supplementary material during any stage of the process, please make sure to provide an updated file. Do not annotate any corrections on a previous version. Please switch off the 'Track Changes' option in Microsoft Office files as these will appear in the published version.

Research data

This journal encourages and enables you to share data that supports your research publication where appropriate, and enables you to interlink the data with your published articles. Research data refers to the results of observations or experimentation that validate research findings. To facilitate reproducibility and data reuse, this journal also encourages you to share your software, code, models, algorithms, protocols, methods and other useful materials related to the project.

Below are a number of ways in which you can associate data with your article or make a statement about the availability of your data when submitting your manuscript. If you are sharing data in one of these ways, you are encouraged to cite the data in your manuscript and reference list. Please refer to the "References" section for more information about data citation. For more information on depositing, sharing and using research data and other relevant research materials, visit the [research data](#) page.

Data linking

If you have made your research data available in a data repository, you can link your article directly to the dataset. Elsevier collaborates with a number of repositories to link articles on ScienceDirect with relevant repositories, giving readers access to underlying data that gives them a better understanding of the research described.

There are different ways to link your datasets to your article. When available, you can directly link your dataset to your article by providing the relevant information in the submission system. For more information, visit the [database linking](#) page.

For [supported data repositories](#) a repository banner will automatically appear next to your published article on ScienceDirect.

In addition, you can link to relevant data or entities through identifiers within the text of your manuscript, using the following format: Database: xxxx (e.g., TAIR: AT1G01020; CCDC: 734053; PDB: 1XFN).

Mendeley Data

This journal supports Mendeley Data, enabling you to deposit any research data (including raw and processed data, video, code, software, algorithms, protocols, and methods) associated with your manuscript in a free-to-use, open access repository. During the submission process, after uploading your manuscript, you will have the opportunity to upload your relevant datasets directly to *Mendeley Data*. The datasets will be listed and directly accessible to readers next to your published article online.

For more information, visit the [Mendeley Data for journals](#) page.

Data statement

To foster transparency, we encourage you to state the availability of your data in your submission. This may be a requirement of your funding body or institution. If your data is unavailable to access or unsuitable to post, you will have the opportunity to indicate why during the submission process, for example by stating that the research data is confidential. The statement will appear with your published article on ScienceDirect. For more information, visit the [Data Statement](#) page.

AFTER ACCEPTANCE

Online proof correction

To ensure a fast publication process of the article, we kindly ask authors to provide us with their proof corrections within two days. Corresponding authors will receive an e-mail with a link to our online proofing system, allowing annotation and correction of proofs online. The environment is similar to MS Word: in addition to editing text, you can also comment on figures/tables and answer questions from the Copy Editor. Web-based proofing provides a faster and less error-prone process by allowing you to directly type your corrections, eliminating the potential introduction of errors.

If preferred, you can still choose to annotate and upload your edits on the PDF version. All instructions for proofing will be given in the e-mail we send to authors, including alternative methods to the online version and PDF.

We will do everything possible to get your article published quickly and accurately. Please use this proof only for checking the typesetting, editing, completeness and correctness of the text, tables and figures. Significant changes to the article as accepted for publication will only be considered at this stage with permission from the Editor. It is important to ensure that all corrections are sent back to us in one communication. Please check carefully before replying, as inclusion of any subsequent corrections cannot be guaranteed. Proofreading is solely your responsibility.

Offprints

The corresponding author will, at no cost, receive a customized [Share Link](#) providing 50 days free access to the final published version of the article on [ScienceDirect](#). The Share Link can be used for sharing the article via any communication channel, including email and social media. For an extra charge, paper offprints can be ordered via the offprint order form which is sent once the article is

accepted for publication. Both corresponding and co-authors may order offprints at any time via Elsevier's [Author Services](#). Corresponding authors who have published their article gold open access do not receive a Share Link as their final published version of the article is available open access on ScienceDirect and can be shared through the article DOI link.

AUTHOR INQUIRIES

Visit the [Elsevier Support Center](#) to find the answers you need. Here you will find everything from Frequently Asked Questions to ways to get in touch.

You can also [check the status of your submitted article](#) or find out [when your accepted article will be published](#).

Appeals

Sometimes editors make mistakes. When we do, we like to hear about them. If an author believes that an editor has made an error in declining a paper, we welcome an appeal. In your appeal letter, which should be sent to tacc@elsevier.com please state why you think the decision is mistaken, and set out your specific responses to any peer reviewers' comments if those seem to have been the main cause of rejection. At least two editors will decide whether to invite a revised manuscript and whether re-review, or otherwise, is indicated.

© Copyright 2018 Elsevier | <https://www.elsevier.com>